

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
 Data File : BE096289.D
 Acq On : 2 May 2018 10:26
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 03 02:03:43 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	75	0.00
2	1,4-Dioxane	0.400	0.355	11.3	72	0.00
3	n-Nitrosodimethylamine	0.400	0.395	1.3	81	0.01
4 S	2-Fluorophenol	0.400	0.362	9.5	72	0.01
5 S	Phenol-d6	0.400	0.377	5.8	74	0.01
6	bis(2-Chloroethyl)ether	0.400	0.367	8.3	72	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	75	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	81	0.00
9	Naphthalene	0.400	0.380	5.0	76	0.00
10	Hexachlorobutadiene	0.400	0.299	25.3#	62	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	75	0.00
12	2-Methylnaphthalene	0.400	0.378	5.5	76	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	75	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.319	20.3	66	0.00
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	77	0.00
16	Acenaphthylene	0.400	0.388	3.0	79	0.00
17	Acenaphthene	0.400	0.375	6.3	76	0.00
18	Fluorene	0.400	0.376	6.0	76	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	73	0.00
20	4-Bromophenyl-phenylether	0.400	0.390	2.5	75	0.00
21	Hexachlorobenzene	0.400	0.394	1.5	74	-0.01
22	Pentachlorophenol	0.400	0.453	-13.2	99	0.00
23	Phenanthrene	0.400	0.391	2.3	75	-0.01
24	Anthracene	0.400	0.372	7.0	73	0.00
25 SURR	Fluoranthene-d10	0.400	0.353	11.8	69	0.00
26	Fluoranthene	0.400	0.363	9.3	71	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	65	0.00
28	Pvrene	0.400	0.049	87.8#	9	0.00
29 S	Terphenyl-d14	0.400	0.365	8.8	65	0.00
30	Benzo(a)anthracene	0.400	0.370	7.5	65	0.00
31	Chrysene	0.400	0.392	2.0	67	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.390	2.5	68	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.349	12.8	60	0.00
34 I	Perylene-d12	0.400	0.400	0.0	63	0.00
35	Benzo(b)fluoranthene	0.400	0.392	2.0	65	0.00
36	Benzo(k)fluoranthene	0.400	0.401	-0.3	67	0.00
37 C	Benzo(a)pyrene	0.400	0.390	2.5	65	0.00
38	Dibenzo(a,h)anthracene	0.400	0.352	12.0	60	0.00
39	Benzo(a,h,i)perylene	0.400	0.368	8.0	61	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
Data File : BE096289.D
Acq On : 2 May 2018 10:26
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 03 02:03:43 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				