

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
 Data File : BE096289.D
 Acq On : 2 May 2018 10:26
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 03 02:03:43 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
2	1,4-Dioxane	0.618	0.548	11.3	72	0.00
3	n-Nitrosodimethylamine	0.757	0.748	1.2	81	0.01
4 S	2-Fluorophenol	1.237	1.119	9.5	72	0.01
5 S	Phenol-d6	1.707	1.608	5.8	74	0.01
6	bis(2-Chloroethyl)ether	1.502	1.380	8.1	72	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
8 S	Nitrobenzene-d5	0.469	0.467	0.4	81	0.00
9	Naphthalene	4.217	4.008	5.0	76	0.00
10	Hexachlorobutadiene	0.187	0.140	25.1#	62	0.00
11 SURR	2-Methylnaphthalene-d10	0.634	0.595	6.2	75	0.00
12	2-Methylnaphthalene	0.698	0.660	5.4	76	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	75	-0.01
14 S	2,4,6-Tribromophenol	0.193	0.154	20.2	66	0.00
15 S	2-Fluorobiphenyl	1.711	1.681	1.8	77	0.00
16	Acenaphthylene	2.136	2.073	2.9	79	0.00
17	Acenaphthene	1.284	1.202	6.4	76	0.00
18	Fluorene	1.588	1.494	5.9	76	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	73	0.00
20	4-Bromophenyl-phenylether	0.238	0.232	2.5	75	0.00
21	Hexachlorobenzene	0.237	0.233	1.7	74	-0.01
22	Pentachlorophenol	0.078	0.087	-11.5	99	0.00
23	Phenanthrene	1.124	1.098	2.3	75	-0.01
24	Anthracene	1.074	1.000	6.9	73	0.00
25 SURR	Fluoranthene-d10	5.430	4.789	11.8	69	0.00
26	Fluoranthene	1.477	1.339	9.3	71	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	65	0.00
28	Pvrene	0.788	0.096	87.8#	9#	0.00
29 S	Terphenyl-d14	0.884	0.806	8.8	65	0.00
30	Benzo(a)anthracene	1.283	1.185	7.6	65	0.00
31	Chrysene	1.239	1.215	1.9	67	-0.01
32	Bis(2-ethylhexyl)phthalate	3.311	3.230	2.4	68	0.00
33	Indeno(1,2,3-cd)pyrene	1.201	1.047	12.8	60	0.00
34 I	Perylene-d12	1.000	1.000	0.0	63	0.00
35	Benzo(b)fluoranthene	1.200	1.175	2.1	65	0.00
36	Benzo(k)fluoranthene	1.234	1.239	-0.4	67	0.00
37 C	Benzo(a)pyrene	1.153	1.123	2.6	65	0.00
38	Dibenzo(a,h)anthracene	1.017	0.895	12.0	60	0.00
39	Benzo(a,h,i)perylene	1.076	0.991	7.9	61	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
Data File : BE096289.D
Acq On : 2 May 2018 10:26
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 03 02:03:43 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			