

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
 Data File : BE096304.D
 Acq On : 2 May 2018 21:20
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 03 01:18:41 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.618	0.534	13.6	82	0.00
3	n-Nitrosodimethylamine	0.757	0.781	-3.2	98	0.00
4 S	2-Fluorophenol	1.237	1.143	7.6	85	0.00
5 S	Phenol-d6	1.707	1.625	4.8	88	0.00
6	bis(2-Chloroethyl)ether	1.502	1.381	8.1	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.469	0.453	3.4	91	0.00
9	Naphthalene	4.217	3.988	5.4	87	-0.01
10	Hexachlorobutadiene	0.187	0.189	-1.1	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.634	0.588	7.3	86	0.00
12	2-Methylnaphthalene	0.698	0.658	5.7	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	-0.01
14 S	2,4,6-Tribromophenol	0.193	0.150	22.3	77	0.00
15 S	2-Fluorobiphenyl	1.711	1.612	5.8	89	0.00
16	Acenaphthylene	2.136	2.018	5.5	92	0.00
17	Acenaphthene	1.284	1.172	8.7	89	0.00
18	Fluorene	1.588	1.516	4.5	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.238	0.225	5.5	88	0.00
21	Hexachlorobenzene	0.237	0.233	1.7	90	-0.01
22	Pentachlorophenol	0.078	0.088	-12.8	123	0.00
23	Phenanthrene	1.124	1.075	4.4	89	-0.01
24	Anthracene	1.074	0.995	7.4	89	0.00
25 SURR	Fluoranthene-d10	5.430	4.846	10.8	85	0.00
26	Fluoranthene	1.477	1.345	8.9	86	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
28	Pvrene	0.788	0.098	87.6#	11#	0.00
29 S	Terphenyl-d14	0.884	0.832	5.9	81	0.00
30	Benzo(a)anthracene	1.283	1.196	6.8	80	0.00
31	Chrysene	1.239	1.194	3.6	79	-0.01
32	Bis(2-ethylhexyl)phthalate	3.311	3.441	-3.9	87	0.00
33	Indeno(1,2,3-cd)pyrene	1.201	1.063	11.5	74	0.00
34 I	Perylene-d12	1.000	1.000	0.0	76	0.00
35	Benzo(b)fluoranthene	1.200	1.178	1.8	78	0.00
36	Benzo(k)fluoranthene	1.234	1.216	1.5	80	0.00
37 C	Benzo(a)pyrene	1.153	1.109	3.8	77	0.00
38	Dibenzo(a,h)anthracene	1.017	0.914	10.1	74	0.00
39	Benzo(a,h,i)perylene	1.076	1.013	5.9	75	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
Data File : BE096304.D
Acq On : 2 May 2018 21:20
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 03 01:18:41 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			