

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
 Data File : BE096315.D
 Acq On : 3 May 2018 4:03
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 03 05:03:28 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.618	0.558	9.7	77	0.00
3	n-Nitrosodimethylamine	0.757	0.827	-9.2	94	0.00
4 S	2-Fluorophenol	1.237	1.126	9.0	76	0.00
5 S	Phenol-d6	1.707	1.607	5.9	78	0.00
6	bis(2-Chloroethyl)ether	1.502	1.371	8.7	75	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
8 S	Nitrobenzene-d5	0.469	0.450	4.1	82	0.00
9	Naphthalene	4.217	3.996	5.2	79	-0.01
10	Hexachlorobutadiene	0.187	0.153	18.2	71	0.00
11 SURR	2-Methylnaphthalene-d10	0.634	0.586	7.6	78	0.00
12	2-Methylnaphthalene	0.698	0.648	7.2	79	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	79	-0.01
14 S	2,4,6-Tribromophenol	0.193	0.158	18.1	71	0.00
15 S	2-Fluorobiphenyl	1.711	1.670	2.4	80	0.00
16	Acenaphthylene	2.136	2.084	2.4	83	0.00
17	Acenaphthene	1.284	1.219	5.1	81	0.00
18	Fluorene	1.588	1.580	0.5	84	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.238	0.214	10.1	82	0.00
21	Hexachlorobenzene	0.237	0.216	8.9	80	-0.01
22	Pentachlorophenol	0.078	0.100	-28.2#	134	0.00
23	Phenanthrene	1.124	1.062	5.5	85	-0.01
24	Anthracene	1.074	0.990	7.8	85	0.00
25 SURR	Fluoranthene-d10	5.430	4.966	8.5	84	0.00
26	Fluoranthene	1.477	1.379	6.6	86	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
28	Pvrene	0.788	0.121	84.6#	13#	0.00
29 S	Terphenyl-d14	0.884	0.819	7.4	80	0.00
30	Benzo(a)anthracene	1.283	1.174	8.5	79	0.00
31	Chrysene	1.239	1.209	2.4	81	-0.01
32	Bis(2-ethylhexyl)phthalate	3.311	3.262	1.5	83	0.00
33	Indeno(1,2,3-cd)pyrene	1.201	1.035	13.8	72	0.00
34 I	Perylene-d12	1.000	1.000	0.0	77	0.00
35	Benzo(b)fluoranthene	1.200	1.200	0.0	81	0.00
36	Benzo(k)fluoranthene	1.234	1.248	-1.1	83	0.00
37 C	Benzo(a)pyrene	1.153	1.119	2.9	79	0.00
38	Dibenzo(a,h)anthracene	1.017	0.839	17.5	69	0.00
39	Benzo(a,h,i)perylene	1.076	1.015	5.7	76	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
Data File : BE096315.D
Acq On : 3 May 2018 4:03
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 03 05:03:28 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			