

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
 Data File : BE096315.D
 Acq On : 3 May 2018 4:03
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 03 05:03:28 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 14:27:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	79	0.00
2	1,4-Dioxane	0.400	0.361	9.8	77	0.00
3	n-Nitrosodimethylamine	0.400	0.437	-9.2	94	0.00
4 S	2-Fluorophenol	0.400	0.364	9.0	76	0.00
5 S	Phenol-d6	0.400	0.377	5.8	78	0.00
6	bis(2-Chloroethyl)ether	0.400	0.365	8.8	75	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	79	0.00
8 S	Nitrobenzene-d5	0.400	0.384	4.0	82	0.00
9	Naphthalene	0.400	0.379	5.3	79	-0.01
10	Hexachlorobutadiene	0.400	0.327	18.3	71	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	78	0.00
12	2-Methylnaphthalene	0.400	0.372	7.0	79	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	79	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.328	18.0	71	0.00
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	80	0.00
16	Acenaphthylene	0.400	0.390	2.5	83	0.00
17	Acenaphthene	0.400	0.380	5.0	81	0.00
18	Fluorene	0.400	0.398	0.5	84	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.400	0.360	10.0	82	0.00
21	Hexachlorobenzene	0.400	0.364	9.0	80	-0.01
22	Pentachlorophenol	0.400	0.501	-25.2#	134	0.00
23	Phenanthrene	0.400	0.378	5.5	85	-0.01
24	Anthracene	0.400	0.369	7.8	85	0.00
25 SURR	Fluoranthene-d10	0.400	0.366	8.5	84	0.00
26	Fluoranthene	0.400	0.373	6.8	86	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	79	0.00
28	Pvrene	0.400	0.062	84.5#	13	0.00
29 S	Terphenyl-d14	0.400	0.371	7.3	80	0.00
30	Benzo(a)anthracene	0.400	0.366	8.5	79	0.00
31	Chrysene	0.400	0.390	2.5	81	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.394	1.5	83	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.345	13.8	72	0.00
34 I	Perylene-d12	0.400	0.400	0.0	77	0.00
35	Benzo(b)fluoranthene	0.400	0.400	0.0	81	0.00
36	Benzo(k)fluoranthene	0.400	0.404	-1.0	83	0.00
37 C	Benzo(a)pyrene	0.400	0.388	3.0	79	0.00
38	Dibenzo(a,h)anthracene	0.400	0.330	17.5	69	0.00
39	Benzo(a,h,i)perylene	0.400	0.377	5.8	76	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE050218\
Data File : BE096315.D
Acq On : 3 May 2018 4:03
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 03 05:03:28 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE041118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 14:27:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			