

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050219\
 Data File : BE099492.D
 Acq On : 2 May 2019 14:30
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 03 04:33:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	-0.01
2	1,4-Dioxane	0.550	0.535	2.7	126	0.00
3	n-Nitrosodimethylamine	0.332	0.404	-21.7	153#	-0.02
4 S	2-Fluorophenol	1.106	1.240	-12.1	131	0.00
5 S	Phenol-d6	1.471	1.675	-13.9	130	-0.01
6	bis(2-Chloroethyl)ether	1.269	1.273	-0.3	114	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.375	0.431	-14.9	122	0.00
9	Naphthalene	4.359	4.564	-4.7	109	0.00
10	Hexachlorobutadiene	0.303	0.320	-5.6	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.736	-2.1	110	0.00
12	2-Methylnaphthalene	0.758	0.777	-2.5	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.252	0.343	-36.1#	154#	0.00
15 S	2-Fluorobiphenyl	1.526	1.575	-3.2	111	-0.01
16	Acenaphthylene	1.957	1.974	-0.9	110	0.00
17	Acenaphthene	1.118	1.127	-0.8	111	0.00
18	Fluorene	1.649	1.715	-4.0	113	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
20	4-Bromophenyl-phenylether	0.233	0.233	0.0	112	0.00
21	Hexachlorobenzene	0.222	0.219	1.4	111	0.00
22	Pentachlorophenol	0.100	0.077	23.0	104	0.00
23	Phenanthrene	1.032	1.040	-0.8	113	0.00
24	Anthracene	0.984	0.999	-1.5	116	0.00
25 SURR	Fluoranthene-d10	5.391	5.664	-5.1	119	0.00
26	Fluoranthene	1.546	1.644	-6.3	120	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
28	Pyrene	1.043	1.242	-19.1	120	0.00
29 S	Terphenyl-d14	0.733	0.865	-18.0	120	0.00
30	Benzo(a)anthracene	1.274	1.445	-13.4	115	0.00
31	Chrysene	1.247	1.420	-13.9	114	0.00
32	Bis(2-ethylhexyl)phthalate	2.216	2.479	-11.9	117	0.00
33	Indeno(1,2,3-cd)pyrene	1.474	1.694	-14.9	115	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.153	1.246	-8.1	119	0.00
36	Benzo(k)fluoranthene	1.112	1.147	-3.1	110	0.00
37 C	Benzo(a)pyrene	1.090	1.163	-6.7	116	0.00
38	Dibenzo(a,h)anthracene	1.115	1.175	-5.4	114	0.00
39	Benzo(g,h,i)perylene	1.118	1.172	-4.8	114	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050219\
Data File : BE099492.D
Acq On : 2 May 2019 14:30
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 03 04:33:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		