

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050219\
 Data File : BE099509.D
 Acq On : 3 May 2019 00:42
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 03 05:30:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	-0.01
2	1,4-Dioxane	0.550	0.515	6.4	124	0.00
3	n-Nitrosodimethylamine	0.332	0.387	-16.6	150#	-0.01
4 S	2-Fluorophenol	1.106	1.178	-6.5	127	0.00
5 S	Phenol-d6	1.471	1.592	-8.2	126	-0.01
6	bis(2-Chloroethyl)ether	1.269	1.250	1.5	114	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
8 S	Nitrobenzene-d5	0.375	0.399	-6.4	123	0.00
9	Naphthalene	4.359	4.471	-2.6	117	0.00
10	Hexachlorobutadiene	0.303	0.312	-3.0	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.732	-1.5	119	0.00
12	2-Methylnaphthalene	0.758	0.772	-1.8	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.252	0.307	-21.8	151#	0.00
15 S	2-Fluorobiphenyl	1.526	1.522	0.3	117	0.00
16	Acenaphthylene	1.957	1.963	-0.3	120	0.00
17	Acenaphthene	1.118	1.128	-0.9	122	0.00
18	Fluorene	1.649	1.640	0.5	118	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
20	4-Bromophenyl-phenylether	0.233	0.221	5.2	117	0.00
21	Hexachlorobenzene	0.222	0.221	0.5	124	0.00
22	Pentachlorophenol	0.100	0.126	-26.0#	187#	0.00
23	Phenanthrene	1.032	1.050	-1.7	125	0.00
24	Anthracene	0.984	1.001	-1.7	128	0.00
25 SURR	Fluoranthene-d10	5.391	5.709	-5.9	132	0.00
26	Fluoranthene	1.546	1.656	-7.1	133	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
28	Pyrene	1.043	1.283	-23.0	132	0.00
29 S	Terphenyl-d14	0.733	0.871	-18.8	129	0.00
30	Benzo(a)anthracene	1.274	1.441	-13.1	122	0.00
31	Chrysene	1.247	1.420	-13.9	122	0.00
32	Bis(2-ethylhexyl)phthalate	2.216	2.338	-5.5	118	0.00
33	Indeno(1,2,3-cd)pyrene	1.474	1.676	-13.7	121	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.153	1.232	-6.9	125	0.00
36	Benzo(k)fluoranthene	1.112	1.148	-3.2	117	0.00
37 C	Benzo(a)pyrene	1.090	1.148	-5.3	122	0.00
38	Dibenzo(a,h)anthracene	1.115	1.163	-4.3	121	0.00
39	Benzo(g,h,i)perylene	1.118	1.150	-2.9	119	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050219\
Data File : BE099509.D
Acq On : 3 May 2019 00:42
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 03 05:30:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			