

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050219\
 Data File : BE099509.D
 Acq On : 3 May 2019 00:42
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 03 05:30:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	-0.01
2	1,4-Dioxane	0.400	0.374	6.5	124	0.00
3	n-Nitrosodimethylamine	0.400	0.467	-16.8	150	-0.01
4 S	2-Fluorophenol	0.400	0.426	-6.5	127	0.00
5 S	Phenol-d6	0.400	0.433	-8.2	126	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	114	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	116	0.00
8 S	Nitrobenzene-d5	0.400	0.426	-6.5	123	0.00
9	Naphthalene	0.400	0.410	-2.5	117	0.00
10	Hexachlorobutadiene	0.400	0.412	-3.0	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.406	-1.5	119	0.00
12	2-Methylnaphthalene	0.400	0.407	-1.7	119	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.487	-21.7	151	0.00
15 S	2-Fluorobiphenyl	0.400	0.399	0.3	117	0.00
16	Acenaphthylene	0.400	0.401	-0.3	120	0.00
17	Acenaphthene	0.400	0.404	-1.0	122	0.00
18	Fluorene	0.400	0.398	0.5	118	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	125	0.00
20	4-Bromophenyl-phenylether	0.400	0.380	5.0	117	0.00
21	Hexachlorobenzene	0.400	0.399	0.3	124	0.00
22	Pentachlorophenol	0.400	0.594	-48.5#	187	0.00
23	Phenanthrene	0.400	0.407	-1.7	125	0.00
24	Anthracene	0.400	0.407	-1.7	128	0.00
25 SURR	Fluoranthene-d10	0.400	0.424	-6.0	132	0.00
26	Fluoranthene	0.400	0.428	-7.0	133	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	122	0.00
28	Pyrene	0.400	0.492	-23.0	132	0.00
29 S	Terphenyl-d14	0.400	0.476	-19.0	129	0.00
30	Benzo(a)anthracene	0.400	0.452	-13.0	122	0.00
31	Chrysene	0.400	0.456	-14.0	122	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.448	-12.0	118	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.455	-13.7	121	0.00
34 I	Perylene-d12	0.400	0.400	0.0	120	0.00
35	Benzo(b)fluoranthene	0.400	0.427	-6.7	125	0.00
36	Benzo(k)fluoranthene	0.400	0.413	-3.2	117	0.00
37 C	Benzo(a)pyrene	0.400	0.421	-5.2	122	0.00
38	Dibenzo(a,h)anthracene	0.400	0.417	-4.2	121	0.00
39	Benzo(g,h,i)perylene	0.400	0.412	-3.0	119	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050219\
Data File : BE099509.D
Acq On : 3 May 2019 00:42
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 03 05:30:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				