

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050316\
 Data File : BE091750.D
 Acq On : 3 May 2016 19:18
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 04 01:40:36 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	151#	0.00
2	1,4-Dioxane	0.752	0.572	23.9	114	0.00
3	n-Nitrosodimethylamine	0.908	0.767	15.5	130	0.02
4 S	2-Fluorophenol	1.228	1.184	3.6	150#	0.00
5 S	Phenol-d6	1.629	1.634	-0.3	161#	0.00
6	bis(2-Chloroethyl)ether	1.433	1.414	1.3	148	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	157#	0.00
8 S	Nitrobenzene-d5	0.323	0.316	2.2	165#	0.00
9	Nitrobenzene	0.338	0.318	5.9	161#	0.00
10	Naphthalene	1.010	1.047	-3.7	161#	0.00
11	Hexachlorobutadiene	0.149	0.163	-9.4	175#	-0.02
12	2-Methylnaphthalene	0.649	0.656	-1.1	159#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	153#	-0.01
14 S	2,4,6-Tribromophenol	0.157	0.149	5.1	160#	0.00
15 S	2-Fluorobiphenyl	1.414	1.411	0.2	155#	0.00
16	Acenaphthylene	1.923	2.055	-6.9	186#	0.00
17	Acenaphthene	1.247	1.238	0.7	153#	0.00
18	Fluorene	1.550	1.564	-0.9	154#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.00
20	4-Bromophenyl-phenylether	0.180	0.191	-6.1	148	0.00
21	Hexachlorobenzene	0.180	0.193	-7.2	151#	0.00
22	Pentachlorophenol	0.085	0.084	1.2	164#	0.00
23	Phenanthrene	1.136	1.168	-2.8	141	0.00
24	Anthracene	1.025	1.118	-9.1	156#	0.00
25	Fluoranthene	1.188	1.372	-15.5	160#	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
27	Benzydine	0.286	0.519	-81.5#	273#	0.00
28	Pvrene	1.130	1.430	-26.5#	159#	0.00
29 S	Terphenyl-d14	0.780	0.935	-19.9	149	0.00
30	Benzo(a)anthracene	1.221	1.326	-8.6	135	0.00
31	3,3'-Dichlorobenzidine	0.623	0.871	-39.8#	196#	-0.02
32	Chrysene	1.269	1.467	-15.6	133	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.906	-144.9#	386#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.342	-18.8	142	0.00
35 I	Perylene-d12	1.000	1.000	0.0	127	0.00
36	Benzo(b)fluoranthene	1.172	1.327	-13.2	148	0.00
37	Benzo(k)fluoranthene	1.191	1.184	0.6	132	0.00
38 C	Benzo(a)pyrene	1.103	1.202	-9.0	146	0.00
39	Dibenzo(a,h)anthracene	1.045	1.114	-6.6	139	-0.01
40	Benzo(a,h,i)perylene	1.062	1.145	-7.8	141	-0.01

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050316\
Data File : BE091750.D
Acq On : 3 May 2016 19:18
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 04 01:40:36 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				