

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050318\
 Data File : BE096333.D
 Acq On : 3 May 2018 21:17
 Operator : SJ/JU
 Sample : J2440-14
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 A3Y29

Manual Integrations
 APPROVED

Sohil
 5/4/2018 1:35:04 PM

Quant Time: May 04 03:00:56 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE050318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 04 01:53:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.35	152	653	0.40	ng/ul	0.00
2) Naphthalene-d8	11.17	136	2816	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.94	164	1811	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.65	188	4667m	0.40	ng/ul	0.00
16) Chrysene-d12	21.74	240	5418m	0.40	ng/ul	0.00
20) Perylene-d12	24.34	264	4156m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.73	152	946	0.19	ng/ul	0.00
14) Fluoranthene-d10	19.63	212	3229m	0.16	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	11.22	128	26262	3.393	ng/ul#	88
5) 2-Methylnaphthalene	12.80	142	3056	0.594	ng/ul	100
7) Acenaphthylene	14.67	152	829	0.088	ng/ul#	86
8) Acenaphthene	15.00	153	4790	0.706	ng/ul	93
9) Fluorene	15.97	166	6375	0.693	ng/ul#	91
12) Phenanthrene	17.69	178	53537m	3.812	ng/ul	
13) Anthracene	17.78	178	8513m	0.627	ng/ul	
15) Fluoranthene	19.66	202	64166m	3.210	ng/ul	
17) Pyrene	20.02	202	15024	2.465	ng/ul#	73
18) Benzo(a)anthracene	21.73	228	12629m	1.492	ng/ul	
19) Chrysene	21.78	228	31850	1.969	ng/ul	98
21) Benzo(b)fluoranthene	23.54	252	38482m	2.706	ng/ul	
22) Benzo(k)fluoranthene	23.59	252	11764m	0.856	ng/ul	
23) Benzo(a)pyrene	24.23	252	26029	1.954	ng/ul#	78
24) Indeno(1,2,3-cd)pyrene	27.11	276	15725	1.066	ng/ul#	73
25) Dibenzo(a,h)anthracene	27.12	278	3882	0.322	ng/ul	97
26) Benzo(a,h,i)perylene	27.97	276	14916	1.201	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050318\
Data File : BE096333.D
Acq On : 3 May 2018 21:17
Operator : SJ/JU
Sample : J2440-14
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
A3Y29

Manual Integrations
APPROVED

Sohil
5/4/2018 1:35:04 PM

Quant Time: May 04 03:00:56 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE050318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 04 01:53:25 2018
Response via : Initial Calibration

