

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050318\
 Data File : BE096335.D
 Acq On : 3 May 2018 22:32
 Operator : SJ/JU
 Sample : J2440-07
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 A3Y25

Manual Integrations
 APPROVED

Sohil
 5/4/2018 1:35:07 PM

Quant Time: May 04 03:53:49 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE050318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 04 01:53:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.35	152	755	0.40	ng/ul	0.00
2) Naphthalene-d8	11.17	136	3390	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.94	164	2160	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.65	188	5365m	0.40	ng/ul	0.00
16) Chrysene-d12	21.74	240	5673m	0.40	ng/ul	0.00
20) Perylene-d12	24.34	264	4750m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.73	152	1619	0.26	ng/ul	0.00
14) Fluoranthene-d10	19.63	212	5690m	0.25	ng/ul	0.00
Target Compounds				Ovalue		
12) Phenanthrene	17.69	178	601m	0.037	ng/ul	
15) Fluoranthene	19.66	202	1701m	0.074	ng/ul	
17) Pyrene	20.02	202	452m	0.071	ng/ul	
18) Benzo(a)anthracene	21.73	228	395m	0.045	ng/ul	
19) Chrysene	21.78	228	886m	0.052	ng/ul	
21) Benzo(b)fluoranthene	23.54	252	1153m	0.071	ng/ul	
22) Benzo(k)fluoranthene	23.59	252	377m	0.024	ng/ul	
23) Benzo(a)pyrene	24.23	252	671	0.044	ng/ul#	83
24) Indeno(1,2,3-cd)pyrene	27.12	276	540	0.032	ng/ul#	77
26) Benzo(g,h,i)perylene	27.99	276	558	0.039	ng/ul#	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050318\
Data File : BE096335.D
Acq On : 3 May 2018 22:32
Operator : SJ/JU
Sample : J2440-07
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
A3Y25

Manual Integrations
APPROVED

Sohil
5/4/2018 1:35:07 PM

Quant Time: May 04 03:53:49 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE050318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 04 01:53:25 2018
Response via : Initial Calibration

