

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050318\
 Data File : BE096326.D
 Acq On : 3 May 2018 17:00
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.477

Quant Time: May 04 01:45:04 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE050318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 03 13:17:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.35	152	559	0.40	ng/ul	0.00
2) Naphthalene-d8	11.17	136	2064	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.94	164	1248	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.65	188	3572	0.40	ng/ul	0.00
16) Chrysene-d12	21.74	240	4667	0.40	ng/ul	0.00
20) Perylene-d12	24.35	264	3663	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.73	152	1560	0.42	ng/ul	0.00
14) Fluoranthene-d10	19.63	212	5744	0.38	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.22	128	2322	0.409	ng/ul#	91
5) 2-Methylnaphthalene	12.80	142	1563	0.414	ng/ul	100
7) Acenaphthylene	14.67	152	2576	0.397	ng/ul	98
8) Acenaphthene	15.00	153	1885	0.403	ng/ul#	93
9) Fluorene	15.97	166	2543	0.401	ng/ul	89
11) Pentachlorophenol	17.31	266	168	0.343	ng/ul	96
12) Phenanthrene	17.69	178	4151	0.386	ng/ul#	90
13) Anthracene	17.78	178	4081	0.393	ng/ul#	92
15) Fluoranthene	19.66	202	5784	0.378	ng/ul	83
17) Pyrene	20.02	202	1607	0.306	ng/ul#	78
18) Benzo(a)anthracene	21.73	228	2663	0.365	ng/ul#	87
19) Chrysene	21.78	228	5326	0.382	ng/ul	96
21) Benzo(b)fluoranthene	23.54	252	4966	0.396	ng/ul	88
22) Benzo(k)fluoranthene	23.59	252	4816	0.398	ng/ul	91
23) Benzo(a)pyrene	24.23	252	4602	0.392	ng/ul#	83
24) Indeno(1,2,3-cd)pyrene	27.12	276	5146	0.396	ng/ul#	79
25) Dibenzo(a,h)anthracene	27.13	278	4185	0.394	ng/ul	93
26) Benzo(g,h,i)perylene	27.98	276	4396	0.402	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050318\
Data File : BE096326.D
Acq On : 3 May 2018 17:00
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.477

Quant Time: May 04 01:45:04 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE050318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 03 13:17:47 2018
Response via : Initial Calibration

