

Data Path : S:\HPCHEM1\BNA_E\DATA\BE050517\
 Data File : BE092942.D
 Acq On : 5 May 2017 22:54
 Operator : SJ/MA
 Sample : I2893-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 3064-PP4

Quant Time: May 06 02:48:37 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 05 14:05:07 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	714	0.40	ng	0.00
7) Naphthalene-d8	10.53	136	2930	0.40	ng	0.00
13) Acenaphthene-d10	14.39	164	1063	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	3581	0.40	ng	0.00
26) Chrysene-d12	21.28	240	3499	0.40	ng	0.00
33) Perylene-d12	23.69	264	774	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.35	112	315	0.20	ng	0.00
5) Phenol-d6	6.94	99	243	0.11	ng	0.00
8) Nitrobenzene-d5	8.90	82	795	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.13	152	1631	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.87	330	102	0.61	ng	0.00
15) 2-Fluorobiphenyl	13.01	172	2616	0.54	ng	0.00
24) Fluoranthene-d10	19.15	212	3436	0.41	ng	0.00
28) Terphenyl-d14	19.73	244	2794	0.46	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Bis(2-ethylhexyl)phthalate	21.20	149	1462	0.65	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA_E\DATA\BE050517\
Data File : BE092942.D
Acq On : 5 May 2017 22:54
Operator : SJ/MA
Sample : I2893-05
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
3064-PP4

Quant Time: May 06 02:48:37 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 05 14:05:07 2017
Response via : Initial Calibration

