

Data Path : S:\HPCHEM1\BNA_E\Data\BE050517\
 Data File : BE092928.D
 Acq On : 5 May 2017 13:33
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE050517

Quant Time: May 05 14:08:13 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 05 14:05:07 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.766	0.737	3.8	93	0.00
3	n-Nitrosodimethylamine	0.594	0.557	6.2	93	0.00
4 S	2-Fluorophenol	0.893	0.867	2.9	98	0.00
5 S	Phenol-d6	1.257	1.190	5.3	97	0.00
6	bis(2-Chloroethyl)ether	1.471	1.386	5.8	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
8 S	Nitrobenzene-d5	0.269	0.271	-0.7	100	0.00
9	Naphthalene	1.069	1.059	0.9	95	0.00
10	Hexachlorobutadiene	0.146	0.145	0.7	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.593	0.584	1.5	95	0.00
12	2-Methylnaphthalene	0.654	0.635	2.9	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
14 S	2,4,6-Tribromophenol	0.063	0.059	6.3	99	0.00
15 S	2-Fluorobiphenyl	1.839	1.851	-0.7	94	0.00
16	Acenaphthylene	6.323	5.972	5.6	96	0.00
17	Acenaphthene	1.263	1.241	1.7	95	-0.02
18	Fluorene	1.560	1.495	4.2	91	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
20	Hexachlorobenzene	0.171	0.163	4.7	94	0.00
21	Pentachlorophenol	0.033	0.029	12.1	97	0.00
22	Phenanthrene	1.126	1.058	6.0	89	0.00
23	Anthracene	0.858	0.795	7.3	89	0.00
24 SURR	Fluoranthene-d10	0.927	0.875	5.6	91	0.00
25	Fluoranthene	4.864	4.530	6.9	90	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
27	Pyrene	5.159	5.209	-1.0	88	0.00
28 S	Terphenyl-d14	0.699	0.728	-4.1	91	0.00
29	Benzo(a)anthracene	0.940	0.912	3.0	89	0.00
30	Chrysene	1.194	1.245	-4.3	91	0.00
31	Bis(2-ethylhexyl)phthalate	0.257	0.249	3.1	84	0.00
32	Indeno(1,2,3-cd)pyrene	4.310	3.900	9.5	80	0.00
33 I	Perylene-d12	1.000	1.000	0.0	71	0.00
34	Benzo(b)fluoranthene	1.309	1.302	0.5	85	0.00
35	Benzo(k)fluoranthene	1.307	1.315	-0.6	80	0.00
36 C	Benzo(a)pyrene	1.069	0.965	9.7	78	-0.01
37	Dibenzo(a,h)anthracene	1.205	1.146	4.9	82	0.00
38	Benzo(g,h,i)perylene	5.114	5.100	0.3	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0