

Data Path : S:\HPCHEM1\BNA_E\Data\BE050517\
 Data File : BE092932.D
 Acq On : 5 May 2017 16:23
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 05 17:03:58 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 05 14:05:07 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.377	5.8	87	0.00
3	n-Nitrosodimethylamine	0.400	0.436	-9.0	103	0.00
4 S	2-Fluorophenol	0.400	0.482	-20.5	116	0.00
5 S	Phenol-d6	0.400	0.470	-17.5	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.388	3.0	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	92	0.00
8 S	Nitrobenzene-d5	0.400	0.472	-18.0	119	0.00
9	Naphthalene	0.400	0.378	5.5	92	0.00
10	Hexachlorobutadiene	0.400	0.391	2.3	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	95	0.00
12	2-Methylnaphthalene	0.400	0.384	4.0	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.400	0.477	-19.2	137	0.00
15 S	2-Fluorobiphenyl	0.400	0.357	10.8	92	0.00
16	Acenaphthylene	0.400	0.400	0.0	111	0.00
17	Acenaphthene	0.400	0.371	7.3	98	0.00
18	Fluorene	0.400	0.364	9.0	94	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	93	0.00
20	Hexachlorobenzene	0.400	0.390	2.5	97	0.00
21	Pentachlorophenol	0.400	0.514	-28.5#	132	0.00
22	Phenanthrene	0.400	0.394	1.5	95	0.00
23	Anthracene	0.400	0.420	-5.0	102	0.00
24 SURR	Fluoranthene-d10	0.400	0.417	-4.2	102	0.00
25	Fluoranthene	0.400	0.413	-3.2	100	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
27	Pyrene	0.400	0.375	6.3	104	0.00
28 S	Terphenyl-d14	0.400	0.350	12.5	97	0.00
29	Benzo(a)anthracene	0.400	0.456	-14.0	133	0.00
30	Chrysene	0.400	0.374	6.5	103	0.00
31	Bis(2-ethylhexyl)phthalate	0.400	0.918	-129.5#	251	0.00
32	Indeno(1,2,3-cd)pyrene	0.400	0.410	-2.5	114	0.00
33 I	Perylene-d12	0.400	0.400	0.0	102	0.00
34	Benzo(b)fluoranthene	0.400	0.370	7.5	113	0.00
35	Benzo(k)fluoranthene	0.400	0.371	7.3	106	0.00
36 C	Benzo(a)pyrene	0.400	0.399	0.3	124	0.00
37	Dibenzo(a,h)anthracene	0.400	0.372	7.0	115	0.00
38	Benzo(g,h,i)perylene	0.400	0.357	10.8	109	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0