

Data Path : S:\HPCHEM1\BNA_E\Data\BE050517\
 Data File : BE092934.D
 Acq On : 5 May 2017 17:59
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 06 01:48:13 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 05 14:05:07 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.766	0.720	6.0	98	0.00
3	n-Nitrosodimethylamine	0.594	0.613	-3.2	111	0.00
4 S	2-Fluorophenol	0.893	1.033	-15.7	126	0.00
5 S	Phenol-d6	1.257	1.377	-9.5	121	0.00
6	bis(2-Chloroethyl)ether	1.471	1.364	7.3	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.269	0.302	-12.3	121	0.00
9	Naphthalene	1.069	1.031	3.6	100	0.00
10	Hexachlorobutadiene	0.146	0.141	3.4	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.593	0.574	3.2	102	0.00
12	2-Methylnaphthalene	0.654	0.635	2.9	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.063	0.074	-17.5	143	0.00
15 S	2-Fluorobiphenyl	1.839	1.698	7.7	99	0.00
16	Acenaphthylene	6.323	6.304	0.3	116	0.00
17	Acenaphthene	1.263	1.194	5.5	105	0.00
18	Fluorene	1.560	1.478	5.3	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	Hexachlorobenzene	0.171	0.171	0.0	105	0.00
21	Pentachlorophenol	0.033	0.038	-15.2	135	0.00
22	Phenanthrene	1.126	1.089	3.3	98	0.00
23	Anthracene	0.858	0.853	0.6	101	0.00
24 SURR	Fluoranthene-d10	0.927	0.919	0.9	102	0.00
25	Fluoranthene	4.864	4.867	-0.1	102	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
27	Pyrene	5.159	4.798	7.0	103	0.00
28 S	Terphenyl-d14	0.699	0.614	12.2	97	0.00
29	Benzo(a)anthracene	0.940	1.017	-8.2	127	0.00
30	Chrysene	1.194	1.144	4.2	106	0.00
31	Bis(2-ethylhexyl)phthalate	0.257	0.346	-34.6#	148	0.00
32	Indeno(1,2,3-cd)pyrene	4.310	3.990	7.4	104	0.00
33 I	Perylene-d12	1.000	1.000	0.0	98	0.00
34	Benzo(b)fluoranthene	1.309	1.217	7.0	110	0.00
35	Benzo(k)fluoranthene	1.307	1.213	7.2	102	0.00
36 C	Benzo(a)pyrene	1.069	0.963	9.9	108	-0.01
37	Dibenzo(a,h)anthracene	1.205	1.073	11.0	106	0.00
38	Benzo(g,h,i)perylene	5.114	4.520	11.6	104	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0