

Data Path : S:\HPCHEM1\BNA_E\Data\BE050517\
 Data File : BE092952.D
 Acq On : 6 May 2017 5:04
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 06 06:54:03 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 05 14:05:07 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	91	0.00
2	1,4-Dioxane	0.400	0.367	8.3	88	0.00
3	n-Nitrosodimethylamine	0.400	0.427	-6.7	105	0.00
4 S	2-Fluorophenol	0.400	0.482	-20.5	120	0.00
5 S	Phenol-d6	0.400	0.477	-19.2	120	0.00
6	bis(2-Chloroethyl)ether	0.400	0.381	4.8	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.466	-16.5	121	0.00
9	Naphthalene	0.400	0.379	5.3	95	0.00
10	Hexachlorobutadiene	0.400	0.360	10.0	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	98	0.00
12	2-Methylnaphthalene	0.400	0.389	2.8	98	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.400	0.552	-38.0#	162	0.00
15 S	2-Fluorobiphenyl	0.400	0.366	8.5	97	0.00
16	Acenaphthylene	0.400	0.406	-1.5	116	0.00
17	Acenaphthene	0.400	0.385	3.8	104	-0.02
18	Fluorene	0.400	0.379	5.3	101	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	101	0.00
20	Hexachlorobenzene	0.400	0.360	10.0	98	0.00
21	Pentachlorophenol	0.400	0.644	-61.0#	195	0.00
22	Phenanthrene	0.400	0.386	3.5	102	0.00
23	Anthracene	0.400	0.402	-0.5	106	0.00
24 SURR	Fluoranthene-d10	0.400	0.403	-0.8	108	0.00
25	Fluoranthene	0.400	0.389	2.8	103	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
27	Pyrene	0.400	0.373	6.8	110	0.00
28 S	Terphenyl-d14	0.400	0.376	6.0	111	0.00
29	Benzo(a)anthracene	0.400	0.402	-0.5	125	0.02
30	Chrysene	0.400	0.380	5.0	112	0.02
31	Bis(2-ethylhexyl)phthalate	0.400	0.535	-33.8#	156	0.00
32	Indeno(1,2,3-cd)pyrene	0.400	0.378	5.5	113	0.00
33 I	Perylene-d12	0.400	0.400	0.0	106	0.00
34	Benzo(b)fluoranthene	0.400	0.353	11.8	112	0.00
35	Benzo(k)fluoranthene	0.400	0.378	5.5	112	0.00
36 C	Benzo(a)pyrene	0.400	0.390	2.5	126	0.00
37	Dibenzo(a,h)anthracene	0.400	0.347	13.3	111	0.00
38	Benzo(g,h,i)perylene	0.400	0.340	15.0	108	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0