

Data Path : Z:\HPCHEM1\BNA E\Data\BE050916\
 Data File : BE091752.D
 Acq On : 9 May 2016 10:54
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 00:59:22 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.752	0.608	19.1	91	0.00
3	n-Nitrosodimethylamine	0.908	0.666	26.7#	85	0.02
4 S	2-Fluorophenol	1.228	1.039	15.4	99	0.00
5 S	Phenol-d6	1.629	1.413	13.3	105	0.00
6	bis(2-Chloroethyl)ether	1.433	1.321	7.8	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
8 S	Nitrobenzene-d5	0.323	0.204	36.8#	80	0.00
9	Nitrobenzene	0.338	0.207	38.8#	79	0.00
10	Naphthalene	1.010	0.983	2.7	114	0.00
11	Hexachlorobutadiene	0.149	0.152	-2.0	124	-0.02
12	2-Methylnaphthalene	0.649	0.626	3.5	115	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	121	-0.01
14 S	2,4,6-Tribromophenol	0.157	0.080	49.0#	68	0.00
15 S	2-Fluorobiphenyl	1.414	1.326	6.2	115	0.00
16	Acenaphthylene	1.923	1.874	2.5	134	0.00
17	Acenaphthene	1.247	1.150	7.8	112	0.00
18	Fluorene	1.550	1.440	7.1	113	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.180	0.172	4.4	107	0.00
21	Hexachlorobenzene	0.180	0.182	-1.1	115	0.00
22	Pentachlorophenol	0.085	0.036	57.6#	57	0.00
23	Phenanthrene	1.136	1.097	3.4	107	0.00
24	Anthracene	1.025	0.979	4.5	110	0.00
25	Fluoranthene	1.188	1.161	2.3	110	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
27	Benzydine	0.286	0.308	-7.7	140	0.00
28	Pvrene	1.130	1.137	-0.6	110	0.00
29 S	Terphenyl-d14	0.780	0.771	1.2	106	0.00
30	Benzo(a)anthracene	1.221	1.166	4.5	103	0.00
31	3,3'-Dichlorobenzidine	0.623	0.655	-5.1	128	-0.02
32	Chrysene	1.269	1.324	-4.3	104	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.391	-5.7	144	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.158	-2.5	106	0.00
35 I	Perylene-d12	1.000	1.000	0.0	108	0.00
36	Benzo(b)fluoranthene	1.172	1.163	0.8	110	0.00
37	Benzo(k)fluoranthene	1.191	1.097	7.9	103	0.00
38 C	Benzo(a)pyrene	1.103	1.070	3.0	110	0.00
39	Dibenzo(a,h)anthracene	1.045	0.999	4.4	106	-0.01
40	Benzo(a,h,i)perylene	1.062	1.046	1.5	110	-0.01

Data Path : Z:\HPCHEM1\BNA E\Data\BE050916\
Data File : BE091752.D
Acq On : 9 May 2016 10:54
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 10 00:59:22 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				