

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050916\
 Data File : BE091757.D
 Acq On : 9 May 2016 14:35
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 01:13:32 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.752	0.599	20.3	87	0.00
3	n-Nitrosodimethylamine	0.908	0.630	30.6#	78	0.02
4 S	2-Fluorophenol	1.228	1.058	13.8	98	0.00
5 S	Phenol-d6	1.629	1.392	14.5	100	0.02
6	bis(2-Chloroethyl)ether	1.433	1.252	12.6	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
8 S	Nitrobenzene-d5	0.323	0.251	22.3	97	0.01
9	Nitrobenzene	0.338	0.251	25.7#	94	0.00
10	Naphthalene	1.010	0.987	2.3	112	0.00
11	Hexachlorobutadiene	0.149	0.148	0.7	117	-0.02
12	2-Methylnaphthalene	0.649	0.621	4.3	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	-0.01
14 S	2,4,6-Tribromophenol	0.157	0.113	28.0#	94	0.00
15 S	2-Fluorobiphenyl	1.414	1.307	7.6	112	0.00
16	Acenaphthylene	1.923	1.857	3.4	132	0.00
17	Acenaphthene	1.247	1.144	8.3	110	0.00
18	Fluorene	1.550	1.416	8.6	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.180	0.171	5.0	105	0.00
21	Hexachlorobenzene	0.180	0.183	-1.7	114	0.00
22	Pentachlorophenol	0.085	0.045	47.1#	71	0.00
23	Phenanthrene	1.136	1.093	3.8	105	0.00
24	Anthracene	1.025	0.957	6.6	107	0.00
25	Fluoranthene	1.188	1.086	8.6	101	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
27	Benzydine	0.286	0.310	-8.4	121	0.00
28	Pvrene	1.130	1.210	-7.1	100	0.00
29 S	Terphenyl-d14	0.780	0.840	-7.7	99	0.00
30	Benzo(a)anthracene	1.221	1.202	1.6	91	0.00
31	3,3'-Dichlorobenzidine	0.623	0.636	-2.1	106	-0.02
32	Chrysene	1.269	1.438	-13.3	96	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.447	-20.8	141	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.224	-8.3	96	0.00
35 I	Perylene-d12	1.000	1.000	0.0	100	0.00
36	Benzo(b)fluoranthene	1.172	1.161	0.9	102	-0.01
37	Benzo(k)fluoranthene	1.191	1.016	14.7	89	0.00
38 C	Benzo(a)pyrene	1.103	1.014	8.1	97	0.00
39	Dibenzo(a,h)anthracene	1.045	0.958	8.3	94	-0.01
40	Benzo(a,h,i)perylene	1.062	1.011	4.8	98	-0.01

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050916\
Data File : BE091757.D
Acq On : 9 May 2016 14:35
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 10 01:13:32 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				