

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050916\
 Data File : BE091757.D
 Acq On : 9 May 2016 14:35
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 01:13:32 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	110	0.00
2	1,4-Dioxane	0.400	0.317	20.8	87	0.00
3	n-Nitrosodimethylamine	0.400	0.277	30.8#	78	0.02
4 S	2-Fluorophenol	0.400	0.345	13.8	98	0.00
5 S	Phenol-d6	0.400	0.342	14.5	100	0.02
6	bis(2-Chloroethyl)ether	0.400	0.349	12.8	95	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	116	0.00
8 S	Nitrobenzene-d5	0.400	0.311	22.3	97	0.01
9	Nitrobenzene	0.400	0.298	25.5#	94	0.00
10	Naphthalene	0.400	0.391	2.3	112	0.00
11	Hexachlorobutadiene	0.400	0.396	1.0	117	-0.02
12	2-Methylnaphthalene	0.400	0.382	4.5	111	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	120	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.398	0.5	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.370	7.5	112	0.00
16	Acenaphthylene	0.400	0.386	3.5	132	0.00
17	Acenaphthene	0.400	0.367	8.3	110	0.00
18	Fluorene	0.400	0.365	8.8	109	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.400	0.380	5.0	105	0.00
21	Hexachlorobenzene	0.400	0.407	-1.7	114	0.00
22	Pentachlorophenol	0.400	0.281	29.8#	71	0.00
23	Phenanthrene	0.400	0.385	3.8	105	0.00
24	Anthracene	0.400	0.374	6.5	107	0.00
25	Fluoranthene	0.400	0.366	8.5	101	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	90	0.00
27	Benzydine	0.400	0.466	-16.5	121	0.00
28	Pvrene	0.400	0.428	-7.0	100	0.00
29 S	Terphenyl-d14	0.400	0.431	-7.7	99	0.00
30	Benzo(a)anthracene	0.400	0.394	1.5	91	0.00
31	3,3'-Dichlorobenzidine	0.400	0.408	-2.0	106	-0.02
32	Chrysene	0.400	0.453	-13.2	96	0.00
33	Bis(2-ethylhexyl)phthalate	0.400	0.483	-20.7	141	-0.02
34	Indeno(1,2,3-cd)pyrene	0.400	0.433	-8.2	96	0.00
35 I	Perylene-d12	5.000	5.000	0.0	100	0.00
36	Benzo(b)fluoranthene	0.400	0.397	0.8	102	-0.01
37	Benzo(k)fluoranthene	0.400	0.341	14.8	89	0.00
38 C	Benzo(a)pyrene	0.400	0.368	8.0	97	0.00
39	Dibenzo(a,h)anthracene	0.400	0.367	8.3	94	-0.01
40	Benzo(a,h,i)perylene	0.400	0.381	4.8	98	-0.01

Data Path : Z:\HPCHEM1\BNA E\DATA\BE050916\
Data File : BE091757.D
Acq On : 9 May 2016 14:35
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 10 01:13:32 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				