

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE050917\
 Data File : BE092964.D
 Acq On : 9 May 2017 14:46
 Operator : SJ/MA
 Sample : I2961-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 OUTFALL001-170502

Manual Integrations
 APPROVED

mohammad
 5/10/2017 11:49:20 AM

Quant Time: May 09 18:43:35 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 09 15:34:16 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	817	0.40	ng	0.00
7) Naphthalene-d8	10.52	136	3001	0.40	ng	0.00
13) Acenaphthene-d10	14.39	164	1599	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	4181	0.40	ng	0.00
26) Chrysene-d12	21.28	240	4660	0.40	ng	0.00
33) Perylene-d12	23.70	264	3861	0.40	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	227	0.12	ng	0.00
5) Phenol-d6	6.95	99	158	0.06	ng	0.00
8) Nitrobenzene-d5	8.90	82	621	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.13	152	1047	0.24	ng	0.00
14) 2,4,6-Tribromophenol	15.87	330	110	0.44	ng	0.00
15) 2-Fluorobiphenyl	13.01	172	1794	0.24	ng	0.00
24) Fluoranthene-d10	19.15	212	3138	0.32	ng	0.00
28) Terphenyl-d14	19.73	244	2721	0.33	ng	0.00
Target Compounds						Qvalue
31) Bis(2-ethylhexyl)phthalate	21.20	149	1431	0.48	ng	# 97
32) Indeno(1,2,3-cd)pyrene	26.24	276	1517	0.03	ng	96
38) Benzo(g,h,i)perylene	27.00	276	1390	0.03	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE050917\
Data File : BE092964.D
Acq On : 9 May 2017 14:46
Operator : SJ/MA
Sample : I2961-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
OUTFALL001-170502

Manual Integrations
APPROVED

mohammad
5/10/2017 11:49:20 AM

Quant Time: May 09 18:43:35 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 09 15:34:16 2017
Response via : Initial Calibration

