

Data Path : Z:\HPCHEM1\BNA_E\Data\BE050917\
 Data File : BE092960.D
 Acq On : 9 May 2017 11:59
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 09 17:12:59 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 09 15:34:16 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.416	-4.0	96	0.00
3	n-Nitrosodimethylamine	0.400	0.498	-24.5	119	0.00
4 S	2-Fluorophenol	0.400	0.530	-32.5#	128	0.00
5 S	Phenol-d6	0.400	0.504	-26.0#	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.440	-10.0	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.452	-13.0	119	0.02
9	Naphthalene	0.400	0.394	1.5	99	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.381	4.8	97	0.00
12	2-Methylnaphthalene	0.400	0.392	2.0	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.400	0.601	-50.2#	184	0.01
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	99	0.01
16	Acenaphthylene	0.400	0.417	-4.2	124	0.00
17	Acenaphthene	0.400	0.398	0.5	112	0.00
18	Fluorene	0.400	0.404	-1.0	112	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	101	0.00
20	Hexachlorobenzene	0.400	0.438	-9.5	119	0.00
21	Pentachlorophenol	0.400	0.559	-39.8#	161	0.02
22	Phenanthrene	0.400	0.424	-6.0	111	0.00
23	Anthracene	0.400	0.461	-15.3	122	0.00
24 SURR	Fluoranthene-d10	0.400	0.469	-17.2	125	0.00
25	Fluoranthene	0.400	0.445	-11.2	118	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	120	0.00
27	Pyrene	0.400	0.379	5.3	118	0.00
28 S	Terphenyl-d14	0.400	0.342	14.5	106	0.00
29	Benzo(a)anthracene	0.400	0.429	-7.2	141	0.02
30	Chrysene	0.400	0.401	-0.3	124	0.02
31	Bis(2-ethylhexyl)phthalate	0.400	0.650	-62.5#	200	0.00
32	Indeno(1,2,3-cd)pyrene	0.400	0.628	-57.0#	197	0.02
33 I	Perylene-d12	0.400	0.400	0.0	125	0.01
34	Benzo(b)fluoranthene	0.400	0.472	-18.0	177	0.01
35	Benzo(k)fluoranthene	0.400	0.456	-14.0	160	0.01
36 C	Benzo(a)pyrene	0.400	0.452	-13.0	173	0.00
37	Dibenzo(a,h)anthracene	0.400	0.494	-23.5	187	0.02
38	Benzo(g,h,i)perylene	0.400	0.558	-39.5#	209	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0