

Data Path : Z:\HPCHEM1\BNA_E\Data\BE050917\
 Data File : BE092960.D
 Acq On : 9 May 2017 11:59
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 09 17:12:59 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 09 15:34:16 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.766	0.797	-4.0	96	0.00
3	n-Nitrosodimethylamine	0.594	0.739	-24.4	119	0.00
4 S	2-Fluorophenol	0.893	1.183	-32.5#	128	0.00
5 S	Phenol-d6	1.257	1.582	-25.9#	123	0.00
6	bis(2-Chloroethyl)ether	1.471	1.619	-10.1	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.269	0.305	-13.4	119	0.02
9	Naphthalene	1.069	1.053	1.5	99	0.00
10	Hexachlorobutadiene	0.146	0.145	0.7	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.593	0.565	4.7	97	0.00
12	2-Methylnaphthalene	0.654	0.642	1.8	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.063	0.094	-49.2#	184#	0.01
15 S	2-Fluorobiphenyl	1.839	1.653	10.1	99	0.01
16	Acenaphthylene	6.323	6.597	-4.3	124	0.00
17	Acenaphthene	1.263	1.256	0.6	112	0.00
18	Fluorene	1.560	1.574	-0.9	112	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	Hexachlorobenzene	0.171	0.187	-9.4	119	0.00
21	Pentachlorophenol	0.033	0.044	-33.3#	161#	0.02
22	Phenanthrene	1.126	1.195	-6.1	111	0.00
23	Anthracene	0.858	0.988	-15.2	122	0.00
24 SURR	Fluoranthene-d10	0.927	1.086	-17.2	125	0.00
25	Fluoranthene	4.864	5.416	-11.3	118	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
27	Pyrene	5.159	4.885	5.3	118	0.00
28 S	Terphenyl-d14	0.699	0.599	14.3	106	0.00
29	Benzo(a)anthracene	0.940	1.008	-7.2	141	0.02
30	Chrysene	1.194	1.197	-0.3	124	0.02
31	Bis(2-ethylhexyl)phthalate	0.257	0.417	-62.3#	200#	0.00
32	Indeno(1,2,3-cd)pyrene	4.310	6.763	-56.9#	197#	0.02
33 I	Perylene-d12	1.000	1.000	0.0	125	0.01
34	Benzo(b)fluoranthene	1.309	1.544	-18.0	177#	0.01
35	Benzo(k)fluoranthene	1.307	1.489	-13.9	160#	0.01
36 C	Benzo(a)pyrene	1.069	1.209	-13.1	173#	0.00
37	Dibenzo(a,h)anthracene	1.205	1.488	-23.5	187#	0.02
38	Benzo(g,h,i)perylene	5.114	7.130	-39.4#	209#	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0