

Data Path : Z:\HPCHEM1\BNA_E\Data\BE050917\
 Data File : BE092965.D
 Acq On : 9 May 2017 15:40
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 09 17:14:35 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 09 15:34:16 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.411	-2.7	104	0.00
3	n-Nitrosodimethylamine	0.400	0.452	-13.0	117	0.00
4 S	2-Fluorophenol	0.400	0.495	-23.7	131	0.00
5 S	Phenol-d6	0.400	0.457	-14.2	122	0.00
6	bis(2-Chloroethyl)ether	0.400	0.406	-1.5	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	0.00
8 S	Nitrobenzene-d5	0.400	0.453	-13.2	121	0.00
9	Naphthalene	0.400	0.389	2.8	100	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.361	9.8	93	0.00
12	2-Methylnaphthalene	0.400	0.370	7.5	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.568	-42.0#	177	0.01
15 S	2-Fluorobiphenyl	0.400	0.332	17.0	93	0.00
16	Acenaphthylene	0.400	0.412	-3.0	125	0.00
17	Acenaphthene	0.400	0.386	3.5	111	0.00
18	Fluorene	0.400	0.392	2.0	111	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	Hexachlorobenzene	0.400	0.468	-17.0	119	0.00
21	Pentachlorophenol	0.400	0.572	-43.0#	155	0.02
22	Phenanthrene	0.400	0.435	-8.7	106	0.00
23	Anthracene	0.400	0.454	-13.5	112	0.00
24 SURR	Fluoranthene-d10	0.400	0.441	-10.2	110	0.00
25	Fluoranthene	0.400	0.447	-11.7	110	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	107	0.00
27	Pyrene	0.400	0.394	1.5	109	0.00
28 S	Terphenyl-d14	0.400	0.365	8.8	101	0.00
29	Benzo(a)anthracene	0.400	0.388	3.0	113	0.02
30	Chrysene	0.400	0.403	-0.8	111	0.02
31	Bis(2-ethylhexyl)phthalate	0.400	0.563	-40.7#	154	0.00
32	Indeno(1,2,3-cd)pyrene	0.400	0.390	2.5	109	0.02
33 I	Perylene-d12	0.400	0.400	0.0	97	0.01
34	Benzo(b)fluoranthene	0.400	0.397	0.8	116	0.00
35	Benzo(k)fluoranthene	0.400	0.375	6.3	102	0.01
36 C	Benzo(a)pyrene	0.400	0.389	2.8	115	0.00
37	Dibenzo(a,h)anthracene	0.400	0.345	13.8	101	0.02
38	Benzo(g,h,i)perylene	0.400	0.382	4.5	111	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0