

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051016\
 Data File : BE091771.D
 Acq On : 11 May 2016 00:15
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 11 01:46:52 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.752	0.557	25.9#	82	0.02
3	n-Nitrosodimethylamine	0.908	0.658	27.5#	83	0.02
4 S	2-Fluorophenol	1.228	1.091	11.2	102	0.00
5 S	Phenol-d6	1.629	1.424	12.6	104	0.02
6	bis(2-Chloroethyl)ether	1.433	1.248	12.9	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.323	0.267	17.3	104	0.01
9	Nitrobenzene	0.338	0.266	21.3	100	0.00
10	Naphthalene	1.010	0.969	4.1	111	0.00
11	Hexachlorobutadiene	0.149	0.148	0.7	119	-0.02
12	2-Methylnaphthalene	0.649	0.609	6.2	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	117	-0.01
14 S	2,4,6-Tribromophenol	0.157	0.125	20.4	103	0.00
15 S	2-Fluorobiphenyl	1.414	1.309	7.4	110	0.00
16	Acenaphthylene	1.923	1.846	4.0	128	0.00
17	Acenaphthene	1.247	1.142	8.4	108	0.00
18	Fluorene	1.550	1.390	10.3	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.180	0.153	15.0	101	0.00
21	Hexachlorobenzene	0.180	0.165	8.3	110	0.00
22	Pentachlorophenol	0.085	0.047	44.7#	78	0.00
23	Phenanthrene	1.136	0.994	12.5	102	0.00
24	Anthracene	1.025	0.852	16.9	101	0.00
25	Fluoranthene	1.188	0.976	17.8	97	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	86	-0.02
27	Benzydine	0.286	0.290	-1.4	107	0.00
28	Pvrene	1.130	1.206	-6.7	95	0.00
29 S	Terphenyl-d14	0.780	0.803	-2.9	90	0.00
30	Benzo(a)anthracene	1.221	1.147	6.1	82	0.00
31	3,3'-Dichlorobenzidine	0.623	0.637	-2.2	101	-0.02
32	Chrysene	1.269	1.302	-2.6	83	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.525	-41.9#	158#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.183	-4.7	88	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	95	-0.01
36	Benzo(b)fluoranthene	1.172	1.040	11.3	87	-0.01
37	Benzo(k)fluoranthene	1.191	1.091	8.4	90	-0.01
38 C	Benzo(a)pyrene	1.103	0.989	10.3	90	-0.01
39	Dibenzo(a,h)anthracene	1.045	0.937	10.3	87	-0.02
40	Benzo(a,h,i)perylene	1.062	0.971	8.6	90	-0.02

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051016\
Data File : BE091771.D
Acq On : 11 May 2016 00:15
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 11 01:46:52 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				