

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051016\
 Data File : BE091771.D
 Acq On : 11 May 2016 00:15
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 11 01:46:52 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	112	0.00
2	1,4-Dioxane	0.400	0.287	28.3#	82	0.02
3	n-Nitrosodimethylamine	0.400	0.290	27.5#	83	0.02
4 S	2-Fluorophenol	0.400	0.355	11.3	102	0.00
5 S	Phenol-d6	0.400	0.350	12.5	104	0.02
6	bis(2-Chloroethyl)ether	0.400	0.348	13.0	96	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.400	0.331	17.3	104	0.01
9	Nitrobenzene	0.400	0.315	21.3	100	0.00
10	Naphthalene	0.400	0.384	4.0	111	0.00
11	Hexachlorobutadiene	0.400	0.397	0.8	119	-0.02
12	2-Methylnaphthalene	0.400	0.375	6.3	110	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	117	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.425	-6.2	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.370	7.5	110	0.00
16	Acenaphthylene	0.400	0.384	4.0	128	0.00
17	Acenaphthene	0.400	0.366	8.5	108	0.00
18	Fluorene	0.400	0.359	10.3	105	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.400	0.340	15.0	101	0.00
21	Hexachlorobenzene	0.400	0.366	8.5	110	0.00
22	Pentachlorophenol	0.400	0.286	28.5#	78	0.00
23	Phenanthrene	0.400	0.350	12.5	102	0.00
24	Anthracene	0.400	0.333	16.8	101	0.00
25	Fluoranthene	0.400	0.329	17.8	97	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	86	-0.02
27	Benzydine	0.400	0.447	-11.7	107	0.00
28	Pvrene	0.400	0.427	-6.7	95	0.00
29 S	Terphenyl-d14	0.400	0.412	-3.0	90	0.00
30	Benzo(a)anthracene	0.400	0.376	6.0	82	0.00
31	3,3'-Dichlorobenzidine	0.400	0.409	-2.2	101	-0.02
32	Chrysene	0.400	0.410	-2.5	83	0.00
33	Bis(2-ethylhexyl)phthalate	0.400	0.568	-42.0#	158	-0.02
34	Indeno(1,2,3-cd)pyrene	0.400	0.419	-4.7	88	-0.01
35 I	Perylene-d12	5.000	5.000	0.0	95	-0.01
36	Benzo(b)fluoranthene	0.400	0.355	11.3	87	-0.01
37	Benzo(k)fluoranthene	0.400	0.366	8.5	90	-0.01
38 C	Benzo(a)pyrene	0.400	0.359	10.3	90	-0.01
39	Dibenzo(a,h)anthracene	0.400	0.359	10.3	87	-0.02
40	Benzo(a,h,i)perylene	0.400	0.366	8.5	90	-0.02

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051016\
Data File : BE091771.D
Acq On : 11 May 2016 00:15
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 11 01:46:52 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				