

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051017\
 Data File : BE092986.D
 Acq On : 10 May 2017 17:40
 Operator : SJ/MA
 Sample : I2893-17RE
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 3076-FBRE

Quant Time: May 11 05:26:26 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 10 03:43:15 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	763	0.40	ng	0.00
7) Naphthalene-d8	10.53	136	3013	0.40	ng	0.00
13) Acenaphthene-d10	14.39	164	542	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	3088	0.40	ng	0.00
26) Chrysene-d12	21.28	240	1517	0.40	ng	0.00
33) Perylene-d12	23.71	264	27	0.40	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	345	0.16	ng	0.00
5) Phenol-d6	6.94	99	434	0.16	ng	0.00
8) Nitrobenzene-d5	8.90	82	842	0.36	ng	0.00
11) 2-Methylnaphthalene-d10	12.13	152	1595	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.87	330	140	0.99	ng	-0.01
15) 2-Fluorobiphenyl	13.01	172	2310	1.14	ng	0.00
24) Fluoranthene-d10	19.15	212	1253	0.16	ng	0.00
28) Terphenyl-d14	19.74	244	2239	0.96	ng	0.00
Target Compounds						Qvalue
22) Phenanthrene	17.18	178	235	0.03	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051017\
Data File : BE092986.D
Acq On : 10 May 2017 17:40
Operator : SJ/MA
Sample : I2893-17RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
3076-FBRE

Quant Time: May 11 05:26:26 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 10 03:43:15 2017
Response via : Initial Calibration

