

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051017\
 Data File : BE092973.D
 Acq On : 9 May 2017 21:02
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE051017

Quant Time: May 10 03:45:11 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 10 03:43:15 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.838	0.810	3.3	98	0.00
3	n-Nitrosodimethylamine	0.692	0.691	0.1	98	0.00
4 S	2-Fluorophenol	1.125	1.077	4.3	95	0.00
5 S	Phenol-d6	1.406	1.341	4.6	97	0.00
6	bis(2-Chloroethyl)ether	1.368	1.304	4.7	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.312	0.305	2.2	95	0.00
9	Naphthalene	1.073	1.040	3.1	93	0.00
10	Hexachlorobutadiene	0.153	0.153	0.0	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.560	0.554	1.1	97	0.00
12	2-Methylnaphthalene	0.616	0.604	1.9	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.104	0.088	15.4	99	0.00
15 S	2-Fluorobiphenyl	1.498	1.452	3.1	96	0.00
16	Acenaphthylene	6.761	6.182	8.6	96	0.00
17	Acenaphthene	1.250	1.172	6.2	96	-0.02
18	Fluorene	1.518	1.439	5.2	95	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	Hexachlorobenzene	0.188	0.178	5.3	97	0.00
21	Pentachlorophenol	0.050	0.040	20.0	88	0.00
22	Phenanthrene	1.137	1.070	5.9	97	0.00
23	Anthracene	0.901	0.818	9.2	93	0.00
24 SURR	Fluoranthene-d10	1.041	0.970	6.8	99	0.00
25	Fluoranthene	5.367	4.836	9.9	95	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
27	Pyrene	5.349	5.430	-1.5	96	0.00
28 S	Terphenyl-d14	0.615	0.593	3.6	91	0.00
29	Benzo(a)anthracene	0.957	0.916	4.3	96	0.00
30	Chrysene	1.279	1.278	0.1	97	0.00
31	Bis(2-ethylhexyl)phthalate	0.429	0.373	13.1	88	0.00
32	Indeno(1,2,3-cd)pyrene	4.419	4.243	4.0	91	-0.02
33 I	Perylene-d12	1.000	1.000	0.0	87	0.00
34	Benzo(b)fluoranthene	1.226	1.172	4.4	98	0.00
35	Benzo(k)fluoranthene	1.313	1.274	3.0	89	-0.01
36 C	Benzo(a)pyrene	1.145	1.084	5.3	95	0.00
37	Dibenzo(a,h)anthracene	1.133	1.053	7.1	89	0.00
38	Benzo(g,h,i)perylene	4.931	4.761	3.4	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0