

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051017\
 Data File : BE092979.D
 Acq On : 10 May 2017 13:21
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 14:08:30 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 10 03:43:15 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.838	0.794	5.3	103	0.00
3	n-Nitrosodimethylamine	0.692	0.688	0.6	104	-0.01
4 S	2-Fluorophenol	1.125	1.048	6.8	99	0.00
5 S	Phenol-d6	1.406	1.318	6.3	102	0.00
6	bis(2-Chloroethyl)ether	1.368	1.387	-1.4	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.312	0.278	10.9	97	0.00
9	Naphthalene	1.073	1.019	5.0	102	0.00
10	Hexachlorobutadiene	0.153	0.145	5.2	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.560	0.538	3.9	105	0.00
12	2-Methylnaphthalene	0.616	0.594	3.6	107	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.104	0.084	19.2	100	0.00
15 S	2-Fluorobiphenyl	1.498	1.486	0.8	105	0.00
16	Acenaphthylene	6.761	6.119	9.5	101	0.00
17	Acenaphthene	1.250	1.198	4.2	104	0.00
18	Fluorene	1.518	1.458	4.0	102	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	Hexachlorobenzene	0.188	0.183	2.7	106	0.00
21	Pentachlorophenol	0.050	0.037	26.0#	88	0.00
22	Phenanthrene	1.137	1.185	-4.2	114	0.00
23	Anthracene	0.901	0.894	0.8	108	0.00
24 SURR	Fluoranthene-d10	1.041	0.975	6.3	105	0.00
25	Fluoranthene	5.367	5.161	3.8	108	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
27	Pyrene	5.349	5.307	0.8	105	0.00
28 S	Terphenyl-d14	0.615	0.584	5.0	101	0.00
29	Benzo(a)anthracene	0.957	0.906	5.3	106	0.00
30	Chrysene	1.279	1.185	7.3	100	0.00
31	Bis(2-ethylhexyl)phthalate	0.429	0.379	11.7	100	0.00
32	Indeno(1,2,3-cd)pyrene	4.419	4.003	9.4	96	0.00
33 I	Perylene-d12	1.000	1.000	0.0	100	0.00
34	Benzo(b)fluoranthene	1.226	1.074	12.4	103	0.00
35	Benzo(k)fluoranthene	1.313	1.235	5.9	99	0.00
36 C	Benzo(a)pyrene	1.145	0.983	14.1	99	0.00
37	Dibenzo(a,h)anthracene	1.133	0.996	12.1	96	0.00
38	Benzo(g,h,i)perylene	4.931	4.411	10.5	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0