

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051017\
 Data File : BE092981.D
 Acq On : 10 May 2017 14:35
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 11 05:09:42 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 10 03:43:15 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.367	8.3	99	0.00
3	n-Nitrosodimethylamine	0.400	0.367	8.3	95	0.00
4 S	2-Fluorophenol	0.400	0.374	6.5	99	0.00
5 S	Phenol-d6	0.400	0.373	6.8	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.369	7.8	96	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	99	0.00
8 S	Nitrobenzene-d5	0.400	0.377	5.8	99	0.00
9	Naphthalene	0.400	0.375	6.3	97	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	103	-0.02
12	2-Methylnaphthalene	0.400	0.387	3.3	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.361	9.8	107	-0.01
15 S	2-Fluorobiphenyl	0.400	0.403	-0.8	101	0.00
16	Acenaphthylene	0.400	0.368	8.0	98	0.00
17	Acenaphthene	0.400	0.389	2.8	101	-0.02
18	Fluorene	0.400	0.392	2.0	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	0.00
20	Hexachlorobenzene	0.400	0.337	15.8	95	0.00
21	Pentachlorophenol	0.400	0.426	-6.5	112	-0.02
22	Phenanthrene	0.400	0.366	8.5	104	0.00
23	Anthracene	0.400	0.347	13.3	98	0.00
24 SURR	Fluoranthene-d10	0.400	0.356	11.0	103	0.00
25	Fluoranthene	0.400	0.345	13.8	100	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	104	0.00
27	Pyrene	0.400	0.378	5.5	102	0.00
28 S	Terphenyl-d14	0.400	0.382	4.5	103	0.00
29	Benzo(a)anthracene	0.400	0.397	0.8	113	0.00
30	Chrysene	0.400	0.371	7.3	102	0.00
31	Bis(2-ethylhexyl)phthalate	0.400	0.422	-5.5	104	0.00
32	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	104	-0.02
33 I	Perylene-d12	0.400	0.400	0.0	105	-0.01
34	Benzo(b)fluoranthene	0.400	0.355	11.3	110	0.00
35	Benzo(k)fluoranthene	0.400	0.364	9.0	101	-0.01
36 C	Benzo(a)pyrene	0.400	0.363	9.3	109	0.00
37	Dibenzo(a,h)anthracene	0.400	0.351	12.3	101	0.00
38	Benzo(g,h,i)perylene	0.400	0.347	13.3	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0