

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051017\
 Data File : BE092987.D
 Acq On : 10 May 2017 18:16
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4EC

Quant Time: May 11 05:32:32 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 10 03:43:15 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.361	9.8	94	0.00
3	n-Nitrosodimethylamine	0.400	0.383	4.3	95	0.00
4 S	2-Fluorophenol	0.400	0.379	5.3	96	0.00
5 S	Phenol-d6	0.400	0.386	3.5	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.379	5.3	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.391	2.3	99	0.00
9	Naphthalene	0.400	0.383	4.3	96	0.00
10	Hexachlorobutadiene	0.400	0.358	10.5	88	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	99	-0.02
12	2-Methylnaphthalene	0.400	0.388	3.0	100	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.384	4.0	109	0.00
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	99	0.00
16	Acenaphthylene	0.400	0.367	8.3	94	0.00
17	Acenaphthene	0.400	0.381	4.8	95	0.00
18	Fluorene	0.400	0.394	1.5	96	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	101	0.00
20	Hexachlorobenzene	0.400	0.369	7.8	100	0.00
21	Pentachlorophenol	0.400	0.433	-8.2	111	0.00
22	Phenanthrene	0.400	0.390	2.5	107	0.00
23	Anthracene	0.400	0.392	2.0	107	0.00
24 SURR	Fluoranthene-d10	0.400	0.396	1.0	111	0.00
25	Fluoranthene	0.400	0.399	0.3	112	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	111	0.00
27	Pyrene	0.400	0.384	4.0	110	0.00
28 S	Terphenyl-d14	0.400	0.386	3.5	111	0.00
29	Benzo(a)anthracene	0.400	0.390	2.5	118	-0.02
30	Chrysene	0.400	0.360	10.0	106	0.00
31	Bis(2-ethylhexyl)phthalate	0.400	0.495	-23.7	127	0.00
32	Indeno(1,2,3-cd)pyrene	0.400	0.346	13.5	100	0.00
33 I	Perylene-d12	0.400	0.400	0.0	104	-0.01
34	Benzo(b)fluoranthene	0.400	0.380	5.0	116	0.00
35	Benzo(k)fluoranthene	0.400	0.401	-0.3	110	-0.01
36 C	Benzo(a)pyrene	0.400	0.363	9.3	108	0.00
37	Dibenzo(a,h)anthracene	0.400	0.333	16.8	95	0.02
38	Benzo(g,h,i)perylene	0.400	0.353	11.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0