

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051117\
 Data File : BE093007.D
 Acq On : 11 May 2017 22:28
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 12 02:20:47 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 11 14:48:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	-0.02
2	1,4-Dioxane	0.400	0.343	14.2	113	0.00
3	n-Nitrosodimethylamine	0.400	0.369	7.8	118	0.00
4 S	2-Fluorophenol	0.400	0.385	3.8	125	0.00
5 S	Phenol-d6	0.400	0.428	-7.0	138	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.357	10.8	114	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	118	0.00
8 S	Nitrobenzene-d5	0.400	0.400	0.0	135	-0.02
9	Naphthalene	0.400	0.357	10.8	117	0.00
10	Hexachlorobutadiene	0.400	0.346	13.5	114	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.365	8.8	122	-0.02
12	2-Methylnaphthalene	0.400	0.366	8.5	123	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	118	0.00
14 S	2,4,6-Tribromophenol	0.400	0.392	2.0	143	0.00
15 S	2-Fluorobiphenyl	0.400	0.376	6.0	127	0.00
16	Acenaphthylene	0.400	0.368	8.0	131	0.00
17	Acenaphthene	0.400	0.359	10.3	121	-0.02
18	Fluorene	0.400	0.365	8.8	123	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	134	0.00
20	4-Bromophenyl-phenylether	0.400	0.330	17.5	125	0.00
21	Hexachlorobenzene	0.400	0.281	29.8#	112	0.00
22	Pentachlorophenol	0.400	0.457	-14.2	156	0.00
23	Phenanthrene	0.400	0.306	23.5	120	0.01
24	Anthracene	0.400	0.338	15.5	123	0.00
25 SURR	Fluoranthene-d10	0.400	0.339	15.3	127	-0.01
26	Fluoranthene	0.400	0.325	18.8	121	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	124	0.00
28	Pvrene	0.400	0.366	8.5	124	0.00
29 S	Terphenyl-d14	0.400	0.374	6.5	138	0.00
30	Benzo(a)anthracene	0.400	0.385	3.8	147	-0.02
31	Chrysene	0.400	0.344	14.0	118	-0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.531	-32.8#	210	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.398	0.5	139	0.00
34 I	Perylene-d12	0.400	0.400	0.0	131	0.00
35	Benzo(b)fluoranthene	0.400	0.332	17.0	117	0.00
36	Benzo(k)fluoranthene	0.400	0.353	11.8	143	0.00
37 C	Benzo(a)pyrene	0.400	0.357	10.8	138	0.00
38	Dibenzo(a,h)anthracene	0.400	0.372	7.0	143	0.00
39	Benzo(a,h,i)perylene	0.400	0.360	10.0	130	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051117\
Data File : BE093007.D
Acq On : 11 May 2017 22:28
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 12 02:20:47 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 11 14:48:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		