


Data Path : Z:\HPCHEM1\BNA E\DATA\BE051117\
 Data File : BE092996.D
 Acq On : 11 May 2017 15:03
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu May 11 14:48:08 2017



AutoFind: Scans 2551, 2552, 2553; Background Corrected with Scan 2539

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	56.7	30639	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	57.1	30861	PASS
70	69	0.00	2	0.7	201	PASS
127	198	10	80	58.7	31754	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	54081	PASS
199	198	5	9	7.4	4006	PASS
275	198	10	60	22.3	12058	PASS
365	198	1	100	2.5	1367	PASS
441	442	0.01	24	15.3	6045	PASS
442	198	50	100	72.8	39389	PASS
443	442	15	24	18.7	7373	PASS