

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051217\
 Data File : BE093009.D
 Acq On : 12 May 2017 9:36
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: May 12 17:21:42 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 11 14:48:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	-0.02
2	1,4-Dioxane	0.802	0.692	13.7	109	0.00
3	n-Nitrosodimethylamine	0.688	0.619	10.0	112	0.00
4 S	2-Fluorophenol	1.097	1.072	2.3	122	0.00
5 S	Phenol-d6	1.376	1.448	-5.2	131	-0.02
6	bis(2-Chloroethyl)ether	1.347	1.186	12.0	109	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.313	0.304	2.9	127	-0.02
9	Naphthalene	1.065	0.933	12.4	110	0.00
10	Hexachlorobutadiene	0.146	0.127	13.0	111	-0.02
11 SURR	2-Methylnaphthalene-d10	0.567	0.512	9.7	116	-0.02
12	2-Methylnaphthalene	0.624	0.572	8.3	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.109	0.107	1.8	138	0.00
15 S	2-Fluorobiphenyl	1.585	1.470	7.3	120	0.00
16	Acenaphthylene	6.788	6.006	11.5	121	0.00
17	Acenaphthene	1.236	1.089	11.9	115	-0.02
18	Fluorene	1.540	1.371	11.0	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.128	0.104	18.8	120	-0.03
21	Hexachlorobenzene	0.191	0.143	25.1#	105	0.00
22	Pentachlorophenol	0.048	0.048	0.0	150	0.00
23	Phenanthrene	0.959	0.732	23.7	117	0.01
24	Anthracene	0.828	0.658	20.5	113	0.00
25 SURR	Fluoranthene-d10	1.059	0.888	16.1	123	-0.02
26	Fluoranthene	5.138	4.187	18.5	119	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	131	0.00
28	Pvrene	5.018	4.315	14.0	123	0.00
29 S	Terphenyl-d14	0.710	0.627	11.7	138	0.00
30	Benzo(a)anthracene	0.983	0.995	-1.2	164#	-0.02
31	Chrysene	1.228	1.045	14.9	123	-0.02
32	Bis(2-ethylhexyl)phthalate	0.434	0.439	-1.2	184#	0.00
33	Indeno(1,2,3-cd)pyrene	3.923	3.762	4.1	142	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	134	-0.01
35	Benzo(b)fluoranthene	1.290	1.111	13.9	125	0.00
36	Benzo(k)fluoranthene	1.289	1.077	16.4	139	-0.01
37 C	Benzo(a)pyrene	1.122	1.013	9.7	143	-0.01
38	Dibenzo(a,h)anthracene	1.016	0.940	7.5	146	0.00
39	Benzo(a,h,i)perylene	4.629	4.081	11.8	130	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051217\
Data File : BE093009.D
Acq On : 12 May 2017 9:36
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 12 17:21:42 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 11 14:48:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			