

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051217\
 Data File : BE093009.D
 Acq On : 12 May 2017 9:36
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 12 17:21:42 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 11 14:48:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	-0.02
2	1,4-Dioxane	0.400	0.345	13.8	109	0.00
3	n-Nitrosodimethylamine	0.400	0.360	10.0	112	0.00
4 S	2-Fluorophenol	0.400	0.391	2.3	122	0.00
5 S	Phenol-d6	0.400	0.421	-5.2	131	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.352	12.0	109	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	114	0.00
8 S	Nitrobenzene-d5	0.400	0.389	2.8	127	-0.02
9	Naphthalene	0.400	0.350	12.5	110	0.00
10	Hexachlorobutadiene	0.400	0.348	13.0	111	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.362	9.5	116	-0.02
12	2-Methylnaphthalene	0.400	0.366	8.5	119	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.400	0.394	1.5	138	0.00
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	120	0.00
16	Acenaphthylene	0.400	0.354	11.5	121	0.00
17	Acenaphthene	0.400	0.353	11.8	115	-0.02
18	Fluorene	0.400	0.356	11.0	115	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.400	0.324	19.0	120	-0.03
21	Hexachlorobenzene	0.400	0.264	34.0#	105	0.00
22	Pentachlorophenol	0.400	0.451	-12.7	150	0.00
23	Phenanthrene	0.400	0.305	23.8	117	0.01
24	Anthracene	0.400	0.318	20.5	113	0.00
25 SURR	Fluoranthene-d10	0.400	0.335	16.3	123	-0.02
26	Fluoranthene	0.400	0.326	18.5	119	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	131	0.00
28	Pvrene	0.400	0.344	14.0	123	0.00
29 S	Terphenyl-d14	0.400	0.353	11.8	138	0.00
30	Benzo(a)anthracene	0.400	0.405	-1.3	164	-0.02
31	Chrysene	0.400	0.341	14.8	123	-0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.459	-14.8	184	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.384	4.0	142	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	134	-0.01
35	Benzo(b)fluoranthene	0.400	0.344	14.0	125	0.00
36	Benzo(k)fluoranthene	0.400	0.334	16.5	139	-0.01
37 C	Benzo(a)pyrene	0.400	0.361	9.8	143	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.370	7.5	146	0.00
39	Benzo(a,h,i)perylene	0.400	0.353	11.8	130	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051217\
Data File : BE093009.D
Acq On : 12 May 2017 9:36
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 12 17:21:42 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 11 14:48:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				