

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051217\
 Data File : BE093023.D
 Acq On : 12 May 2017 18:15
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4EC

Quant Time: May 12 23:52:22 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 11 14:48:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	-0.02
2	1,4-Dioxane	0.802	0.748	6.7	110	0.00
3	n-Nitrosodimethylamine	0.688	0.746	-8.4	125	0.00
4 S	2-Fluorophenol	1.097	1.252	-14.1	132	0.00
5 S	Phenol-d6	1.376	1.868	-35.8#	156#	-0.02
6	bis(2-Chloroethyl)ether	1.347	1.374	-2.0	117	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.313	0.377	-20.4	155#	-0.02
9	Naphthalene	1.065	1.022	4.0	119	0.00
10	Hexachlorobutadiene	0.146	0.138	5.5	119	-0.02
11 SURR	2-Methylnaphthalene-d10	0.567	0.574	-1.2	129	-0.02
12	2-Methylnaphthalene	0.624	0.646	-3.5	133	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.109	0.149	-36.7#	196#	0.00
15 S	2-Fluorobiphenyl	1.585	1.610	-1.6	134	0.00
16	Acenaphthylene	6.788	7.272	-7.1	150#	0.00
17	Acenaphthene	1.236	1.235	0.1	133	-0.02
18	Fluorene	1.540	1.554	-0.9	133	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	132	0.00
20	4-Bromophenyl-phenylether	0.128	0.116	9.4	135	-0.03
21	Hexachlorobenzene	0.191	0.156	18.3	116	0.00
22	Pentachlorophenol	0.048	0.070	-45.8#	222#	0.00
23	Phenanthrene	0.959	0.819	14.6	132	0.01
24	Anthracene	0.828	0.774	6.5	134	0.00
25 SURR	Fluoranthene-d10	1.059	1.062	-0.3	148	-0.02
26	Fluoranthene	5.138	4.764	7.3	136	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
28	Pvrene	5.018	4.916	2.0	139	0.00
29 S	Terphenyl-d14	0.710	0.734	-3.4	159#	0.00
30	Benzo(a)anthracene	0.983	1.130	-15.0	183#	-0.02
31	Chrysene	1.228	1.117	9.0	130	-0.02
32	Bis(2-ethylhexyl)phthalate	0.434	1.103	-154.1#	457#	0.00
33	Indeno(1,2,3-cd)pyrene	3.923	4.500	-14.7	168#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	142	-0.01
35	Benzo(b)fluoranthene	1.290	1.178	8.7	141	0.00
36	Benzo(k)fluoranthene	1.289	1.219	5.4	167#	-0.01
37 C	Benzo(a)pyrene	1.122	1.111	1.0	166#	0.00
38	Dibenzo(a,h)anthracene	1.016	1.071	-5.4	176#	0.00
39	Benzo(a,h,i)perylene	4.629	4.421	4.5	150	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051217\
Data File : BE093023.D
Acq On : 12 May 2017 18:15
Operator : SJ/MA
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4EC

Quant Time: May 12 23:52:22 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 11 14:48:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			