

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051217\
 Data File : BE093023.D
 Acq On : 12 May 2017 18:15
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4EC

Quant Time: May 12 23:52:22 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 11 14:48:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	-0.02
2	1,4-Dioxane	0.400	0.373	6.8	110	0.00
3	n-Nitrosodimethylamine	0.400	0.434	-8.5	125	0.00
4 S	2-Fluorophenol	0.400	0.456	-14.0	132	0.00
5 S	Phenol-d6	0.400	0.543	-35.8#	156	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.408	-2.0	117	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.481	-20.2	155	-0.02
9	Naphthalene	0.400	0.384	4.0	119	0.00
10	Hexachlorobutadiene	0.400	0.377	5.8	119	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	129	-0.02
12	2-Methylnaphthalene	0.400	0.414	-3.5	133	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.400	0.548	-37.0#	196	0.00
15 S	2-Fluorobiphenyl	0.400	0.406	-1.5	134	0.00
16	Acenaphthylene	0.400	0.429	-7.2	150	0.00
17	Acenaphthene	0.400	0.400	0.0	133	-0.02
18	Fluorene	0.400	0.404	-1.0	133	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	132	0.00
20	4-Bromophenyl-phenylether	0.400	0.361	9.8	135	-0.03
21	Hexachlorobenzene	0.400	0.299	25.3#	116	0.00
22	Pentachlorophenol	0.400	0.628	-57.0#	222	0.00
23	Phenanthrene	0.400	0.353	11.8	132	0.01
24	Anthracene	0.400	0.374	6.5	134	0.00
25 SURR	Fluoranthene-d10	0.400	0.401	-0.3	148	-0.02
26	Fluoranthene	0.400	0.371	7.3	136	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	129	0.00
28	Pvrene	0.400	0.392	2.0	139	0.00
29 S	Terphenyl-d14	0.400	0.413	-3.2	159	0.00
30	Benzo(a)anthracene	0.400	0.460	-15.0	183	-0.02
31	Chrysene	0.400	0.364	9.0	130	-0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.994	-148.5#	457	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.459	-14.8	168	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	142	-0.01
35	Benzo(b)fluoranthene	0.400	0.365	8.8	141	0.00
36	Benzo(k)fluoranthene	0.400	0.378	5.5	167	-0.01
37 C	Benzo(a)pyrene	0.400	0.396	1.0	166	0.00
38	Dibenzo(a,h)anthracene	0.400	0.422	-5.5	176	0.00
39	Benzo(a,h,i)perylene	0.400	0.382	4.5	150	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051217\
Data File : BE093023.D
Acq On : 12 May 2017 18:15
Operator : SJ/MA
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4EC

Quant Time: May 12 23:52:22 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 11 14:48:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				