

Data Path : Z:\HPCHEM1\BNA E\Data\BE051517\
 Data File : BE093031.D
 Acq On : 15 May 2017 12:59
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICV0.4

Quant Time: May 15 13:49:21 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 15 13:35:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.810	0.775	4.3	102	0.00
3	n-Nitrosodimethylamine	0.724	0.708	2.2	101	-0.01
4 S	2-Fluorophenol	1.226	1.181	3.7	98	0.00
5 S	Phenol-d6	1.656	1.622	2.1	101	0.00
6	bis(2-Chloroethyl)ether	1.349	1.342	0.5	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
8 S	Nitrobenzene-d5	0.350	0.339	3.1	102	0.00
9	Naphthalene	1.054	1.058	-0.4	102	0.00
10	Hexachlorobutadiene	0.138	0.138	0.0	103	0.02
11 SURR	2-Methylnaphthalene-d10	0.586	0.586	0.0	102	0.02
12	2-Methylnaphthalene	0.653	0.655	-0.3	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.118	0.107	9.3	96	0.00
15 S	2-Fluorobiphenyl	1.645	1.656	-0.7	103	0.00
16	Acenaphthylene	7.015	6.692	4.6	103	0.00
17	Acenaphthene	1.263	1.265	-0.2	105	0.02
18	Fluorene	1.545	1.522	1.5	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.00
20	4-Bromophenyl-phenylether	0.124	0.129	-4.0	100	0.03
21	Hexachlorobenzene	0.163	0.204	-25.2#	120	0.00
22	Pentachlorophenol	0.050	0.046	8.0	105	0.00
23	Phenanthrene	0.867	0.947	-9.2	107	0.03
24	Anthracene	0.816	0.930	-14.0	120	0.00
25 SURR	Fluoranthene-d10	1.051	1.017	3.2	92	0.00
26	Fluoranthene	4.922	5.009	-1.8	99	0.02
27 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
28	Pvrene	4.819	4.856	-0.8	95	0.00
29 S	Terphenyl-d14	0.747	0.745	0.3	95	0.00
30	Benzo(a)anthracene	1.088	1.025	5.8	91	0.00
31	Chrysene	1.163	1.159	0.3	97	0.02
32	Bis(2-ethylhexyl)phthalate	0.627	0.440	29.8#	86	0.00
33	Indeno(1,2,3-cd)pyrene	4.256	4.426	-4.0	102	0.02
34 I	Perylene-d12	1.000	1.000	0.0	85	-0.01
35	Benzo(b)fluoranthene	1.264	1.245	1.5	98	0.00
36	Benzo(k)fluoranthene	1.268	1.290	-1.7	98	0.00
37 C	Benzo(a)pyrene	1.157	1.105	4.5	95	0.00
38	Dibenzo(a,h)anthracene	1.076	1.148	-6.7	105	0.02
39	Benzo(a,h,i)perylene	4.686	4.732	-1.0	98	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE051517\
Data File : BE093031.D
Acq On : 15 May 2017 12:59
Operator : SJ/MA
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICV0.4

Quant Time: May 15 13:49:21 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 15 13:35:57 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			