

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051517\
 Data File : BE093033.D
 Acq On : 15 May 2017 14:23
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 16 01:10:38 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 15 13:35:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.810	0.694	14.3	101	0.00
3	n-Nitrosodimethylamine	0.724	0.649	10.4	102	0.00
4 S	2-Fluorophenol	1.226	1.079	12.0	99	0.00
5 S	Phenol-d6	1.656	1.480	10.6	102	0.00
6	bis(2-Chloroethyl)ether	1.349	1.223	9.3	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.350	0.300	14.3	102	0.00
9	Naphthalene	1.054	0.934	11.4	102	0.00
10	Hexachlorobutadiene	0.138	0.124	10.1	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.586	0.522	10.9	103	0.00
12	2-Methylnaphthalene	0.653	0.577	11.6	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.118	0.095	19.5	97	0.00
15 S	2-Fluorobiphenyl	1.645	1.456	11.5	102	0.00
16	Acenaphthylene	7.015	5.846	16.7	102	0.00
17	Acenaphthene	1.263	1.112	12.0	104	0.00
18	Fluorene	1.545	1.341	13.2	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.124	0.104	16.1	98	0.00
21	Hexachlorobenzene	0.163	0.154	5.5	111	0.00
22	Pentachlorophenol	0.050	0.053	-6.0	148	0.00
23	Phenanthrene	0.867	0.712	17.9	99	0.03
24	Anthracene	0.816	0.693	15.1	109	0.00
25 SURR	Fluoranthene-d10	1.051	0.884	15.9	97	0.00
26	Fluoranthene	4.922	4.113	16.4	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	4.819	4.314	10.5	101	0.00
29 S	Terphenyl-d14	0.747	0.666	10.8	101	0.00
30	Benzo(a)anthracene	1.088	0.971	10.8	103	0.00
31	Chrysene	1.163	1.025	11.9	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.627	0.416	33.7#	98	0.00
33	Indeno(1,2,3-cd)pyrene	4.256	3.656	14.1	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	99	-0.01
35	Benzo(b)fluoranthene	1.264	1.083	14.3	99	0.00
36	Benzo(k)fluoranthene	1.268	1.101	13.2	97	0.00
37 C	Benzo(a)pyrene	1.157	1.019	11.9	102	0.00
38	Dibenzo(a,h)anthracene	1.076	0.945	12.2	100	0.00
39	Benzo(a,h,i)perylene	4.686	4.059	13.4	98	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051517\
Data File : BE093033.D
Acq On : 15 May 2017 14:23
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 16 01:10:38 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 15 13:35:57 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			