

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051517\
 Data File : BE093044.D
 Acq On : 15 May 2017 21:08
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 16 01:14:22 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 15 13:35:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.810	0.669	17.4	97	0.00
3	n-Nitrosodimethylamine	0.724	0.650	10.2	102	0.00
4 S	2-Fluorophenol	1.226	1.096	10.6	100	0.00
5 S	Phenol-d6	1.656	1.515	8.5	103	0.00
6	bis(2-Chloroethyl)ether	1.349	1.204	10.7	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.350	0.320	8.6	109	0.00
9	Naphthalene	1.054	0.939	10.9	102	-0.02
10	Hexachlorobutadiene	0.138	0.125	9.4	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.586	0.516	11.9	101	0.00
12	2-Methylnaphthalene	0.653	0.586	10.3	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.118	0.114	3.4	116	-0.01
15 S	2-Fluorobiphenyl	1.645	1.483	9.8	104	0.00
16	Acenaphthylene	7.015	6.214	11.4	108	0.00
17	Acenaphthene	1.263	1.114	11.8	104	0.00
18	Fluorene	1.545	1.376	10.9	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4-Bromophenyl-phenylether	0.124	0.108	12.9	115	0.00
21	Hexachlorobenzene	0.163	0.124	23.9	101	0.00
22	Pentachlorophenol	0.050	0.045	10.0	143	0.00
23	Phenanthrene	0.867	0.691	20.3	108	0.00
24	Anthracene	0.816	0.598	26.7#	106	0.00
25 SURR	Fluoranthene-d10	1.051	0.872	17.0	108	0.00
26	Fluoranthene	4.922	4.014	18.4	109	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	4.819	4.491	6.8	106	0.00
29 S	Terphenyl-d14	0.747	0.678	9.2	104	0.00
30	Benzo(a)anthracene	1.088	1.009	7.3	108	0.00
31	Chrysene	1.163	1.050	9.7	106	0.00
32	Bis(2-ethylhexyl)phthalate	0.627	0.547	12.8	129	0.00
33	Indeno(1,2,3-cd)pyrene	4.256	3.838	9.8	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	-0.01
35	Benzo(b)fluoranthene	1.264	1.093	13.5	106	0.00
36	Benzo(k)fluoranthene	1.268	1.099	13.3	103	0.00
37 C	Benzo(a)pyrene	1.157	0.998	13.7	106	0.00
38	Dibenzo(a,h)anthracene	1.076	0.961	10.7	108	0.00
39	Benzo(a,h,i)perylene	4.686	4.026	14.1	103	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051517\
Data File : BE093044.D
Acq On : 15 May 2017 21:08
Operator : SJ/MA
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 16 01:14:22 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 15 13:35:57 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			