

Data Path : Z:\HPCHEM1\BNA E\Data\BE051616\
 Data File : BE091773.D
 Acq On : 16 May 2016 14:41
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 17 04:20:02 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.752	0.577	23.3	95	0.02
3	n-Nitrosodimethylamine	0.908	0.656	27.8#	92	0.02
4 S	2-Fluorophenol	1.228	1.050	14.5	110	0.00
5 S	Phenol-d6	1.629	1.380	15.3	113	0.00
6	bis(2-Chloroethyl)ether	1.433	1.297	9.5	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	135	0.00
8 S	Nitrobenzene-d5	0.323	0.242	25.1#	109	0.00
9	Nitrobenzene	0.338	0.245	27.5#	106	0.00
10	Naphthalene	1.010	0.985	2.5	130	0.00
11	Hexachlorobutadiene	0.149	0.143	4.0	132	-0.02
12	2-Methylnaphthalene	0.649	0.614	5.4	128	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	139	-0.01
14 S	2,4,6-Tribromophenol	0.157	0.117	25.5#	113	0.00
15 S	2-Fluorobiphenyl	1.414	1.312	7.2	130	0.00
16	Acenaphthylene	1.923	1.759	8.5	144	0.00
17	Acenaphthene	1.247	1.151	7.7	128	0.00
18	Fluorene	1.550	1.409	9.1	126	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
20	4-Bromophenyl-phenylether	0.180	0.164	8.9	117	0.00
21	Hexachlorobenzene	0.180	0.180	0.0	130	0.00
22	Pentachlorophenol	0.085	0.072	15.3	130	0.00
23	Phenanthrene	1.136	1.096	3.5	122	0.00
24	Anthracene	1.025	0.930	9.3	120	0.00
25	Fluoranthene	1.188	1.035	12.9	112	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
27	Benzydine	0.286	0.294	-2.8	126	0.00
28	Pvrene	1.130	1.235	-9.3	112	0.00
29 S	Terphenyl-d14	0.780	0.822	-5.4	107	0.00
30	Benzo(a)anthracene	1.221	1.137	6.9	94	0.00
31	3,3'-Dichlorobenzidine	0.623	0.521	16.4	95	-0.02
32	Chrysene	1.269	1.387	-9.3	102	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.360	2.7	125	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.190	-5.3	102	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	112	0.00
36	Benzo(b)fluoranthene	1.172	1.103	5.9	108	0.00
37	Benzo(k)fluoranthene	1.191	1.021	14.3	100	0.00
38 C	Benzo(a)pyrene	1.103	0.942	14.6	101	0.00
39	Dibenzo(a,h)anthracene	1.045	0.906	13.3	100	-0.01
40	Benzo(a,h,i)perylene	1.062	0.937	11.8	102	-0.02

Data Path : Z:\HPCHEM1\BNA E\Data\BE051616\
Data File : BE091773.D
Acq On : 16 May 2016 14:41
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 17 04:20:02 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				