

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051616\
 Data File : BE091788.D
 Acq On : 17 May 2016 00:40
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 17 05:11:43 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.752	0.572	23.9	94	0.00
3	n-Nitrosodimethylamine	0.908	0.661	27.2#	93	0.02
4 S	2-Fluorophenol	1.228	1.100	10.4	115	0.00
5 S	Phenol-d6	1.629	1.462	10.3	119	0.00
6	bis(2-Chloroethyl)ether	1.433	1.293	9.8	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	136	0.00
8 S	Nitrobenzene-d5	0.323	0.261	19.2	118	0.00
9	Nitrobenzene	0.338	0.260	23.1	114	0.00
10	Naphthalene	1.010	0.966	4.4	129	0.00
11	Hexachlorobutadiene	0.149	0.143	4.0	133	-0.02
12	2-Methylnaphthalene	0.649	0.623	4.0	131	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	140	-0.01
14 S	2,4,6-Tribromophenol	0.157	0.124	21.0	121	0.00
15 S	2-Fluorobiphenyl	1.414	1.301	8.0	131	0.00
16	Acenaphthylene	1.923	1.813	5.7	150#	0.00
17	Acenaphthene	1.247	1.157	7.2	131	0.00
18	Fluorene	1.550	1.418	8.5	128	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
20	4-Bromophenyl-phenylether	0.180	0.172	4.4	122	0.00
21	Hexachlorobenzene	0.180	0.182	-1.1	130	0.00
22	Pentachlorophenol	0.085	0.063	25.9#	113	0.00
23	Phenanthrene	1.136	1.105	2.7	122	0.00
24	Anthracene	1.025	0.949	7.4	121	0.00
25	Fluoranthene	1.188	1.127	5.1	121	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	104	-0.02
27	Benzidine	0.286	0.256	10.5	115	0.00
28	Pvrene	1.130	1.209	-7.0	115	0.00
29 S	Terphenyl-d14	0.780	0.783	-0.4	106	0.00
30	Benzo(a)anthracene	1.221	1.189	2.6	103	0.00
31	3,3'-Dichlorobenzidine	0.623	0.652	-4.7	125	-0.02
32	Chrysene	1.269	1.252	1.3	96	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.517	-39.7#	188#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.217	-7.7	109	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	115	-0.01
36	Benzo(b)fluoranthene	1.172	1.058	9.7	107	-0.01
37	Benzo(k)fluoranthene	1.191	1.111	6.7	112	-0.01
38 C	Benzo(a)pyrene	1.103	1.003	9.1	110	-0.01
39	Dibenzo(a,h)anthracene	1.045	0.966	7.6	109	-0.02
40	Benzo(a,h,i)perylene	1.062	0.999	5.9	111	-0.02

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051616\
Data File : BE091788.D
Acq On : 17 May 2016 00:40
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 17 05:11:43 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				