

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051617\
 Data File : BE093047.D
 Acq On : 16 May 2017 10:28
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 18 00:53:08 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 15 13:35:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.810	0.706	12.8	102	0.00
3	n-Nitrosodimethylamine	0.724	0.663	8.4	104	-0.01
4 S	2-Fluorophenol	1.226	1.053	14.1	96	0.00
5 S	Phenol-d6	1.656	1.507	9.0	103	0.00
6	bis(2-Chloroethyl)ether	1.349	1.236	8.4	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.350	0.304	13.1	101	0.00
9	Naphthalene	1.054	0.931	11.7	99	0.00
10	Hexachlorobutadiene	0.138	0.125	9.4	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.586	0.524	10.6	100	0.01
12	2-Methylnaphthalene	0.653	0.580	11.2	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.118	0.109	7.6	104	0.00
15 S	2-Fluorobiphenyl	1.645	1.546	6.0	102	0.00
16	Acenaphthylene	7.015	6.142	12.4	100	0.00
17	Acenaphthene	1.263	1.129	10.6	99	0.02
18	Fluorene	1.545	1.406	9.0	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.124	0.101	18.5	101	0.00
21	Hexachlorobenzene	0.163	0.142	12.9	108	0.00
22	Pentachlorophenol	0.050	0.044	12.0	131	0.00
23	Phenanthrene	0.867	0.666	23.2	98	0.03
24	Anthracene	0.816	0.649	20.5	108	0.00
25 SURR	Fluoranthene-d10	1.051	0.808	23.1	94	0.00
26	Fluoranthene	4.922	3.881	21.1	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pvrene	4.819	4.606	4.4	102	0.00
29 S	Terphenyl-d14	0.747	0.703	5.9	102	0.00
30	Benzo(a)anthracene	1.088	0.973	10.6	99	0.00
31	Chrysene	1.163	1.048	9.9	100	0.00
32	Bis(2-ethylhexyl)phthalate	0.627	0.459	26.8#	102	0.00
33	Indeno(1,2,3-cd)pyrene	4.256	4.007	5.9	105	0.00
34 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
35	Benzo(b)fluoranthene	1.264	1.099	13.1	103	0.00
36	Benzo(k)fluoranthene	1.268	1.088	14.2	98	0.00
37 C	Benzo(a)pyrene	1.157	1.005	13.1	103	0.00
38	Dibenzo(a,h)anthracene	1.076	0.970	9.9	105	0.00
39	Benzo(a,h,i)perylene	4.686	4.097	12.6	101	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051617\
Data File : BE093047.D
Acq On : 16 May 2017 10:28
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 18 00:53:08 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 15 13:35:57 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			