

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051617\
 Data File : BE093051.D
 Acq On : 16 May 2017 13:27
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 18 00:54:27 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 15 13:35:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.810	0.730	9.9	136	0.00
3	n-Nitrosodimethylamine	0.724	0.822	-13.5	167#	-0.01
4 S	2-Fluorophenol	1.226	1.159	5.5	137	0.00
5 S	Phenol-d6	1.656	1.813	-9.5	160#	0.00
6	bis(2-Chloroethyl)ether	1.349	1.362	-1.0	148	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	139	0.00
8 S	Nitrobenzene-d5	0.350	0.392	-12.0	181#	0.00
9	Naphthalene	1.054	1.039	1.4	153#	-0.02
10	Hexachlorobutadiene	0.138	0.132	4.3	150#	0.00
11 SURR	2-Methylnaphthalene-d10	0.586	0.530	9.6	141	0.00
12	2-Methylnaphthalene	0.653	0.577	11.6	139	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
14 S	2,4,6-Tribromophenol	0.118	0.161	-36.4#	212#	-0.01
15 S	2-Fluorobiphenyl	1.645	1.564	4.9	142	0.00
16	Acenaphthylene	7.015	7.229	-3.1	163#	0.00
17	Acenaphthene	1.263	1.211	4.1	147	0.00
18	Fluorene	1.545	1.536	0.6	151#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
20	4-Bromophenyl-phenylether	0.124	0.126	-1.6	176#	0.00
21	Hexachlorobenzene	0.163	0.128	21.5	137	0.00
22	Pentachlorophenol	0.050	0.070	-40.0#	290#	-0.03
23	Phenanthrene	0.867	0.759	12.5	156#	0.00
24	Anthracene	0.816	0.687	15.8	160#	0.00
25 SURR	Fluoranthene-d10	1.051	1.010	3.9	164#	0.00
26	Fluoranthene	4.922	4.451	9.6	159#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	152#	-0.02
28	Pvrene	4.819	4.570	5.2	157#	0.00
29 S	Terphenyl-d14	0.747	0.710	5.0	159#	0.00
30	Benzo(a)anthracene	1.088	1.159	-6.5	182#	0.00
31	Chrysene	1.163	1.113	4.3	164#	0.00
32	Bis(2-ethylhexyl)phthalate	0.627	1.055	-68.3#	365#	0.00
33	Indeno(1,2,3-cd)pyrene	4.256	4.418	-3.8	179#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	157#	-0.01
35	Benzo(b)fluoranthene	1.264	1.246	1.4	181#	0.00
36	Benzo(k)fluoranthene	1.268	1.150	9.3	162#	0.00
37 C	Benzo(a)pyrene	1.157	1.113	3.8	177#	-0.01
38	Dibenzo(a,h)anthracene	1.076	1.102	-2.4	186#	0.00
39	Benzo(a,h,i)perylene	4.686	4.504	3.9	173#	-0.02

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051617\
Data File : BE093051.D
Acq On : 16 May 2017 13:27
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 18 00:54:27 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 15 13:35:57 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			