

Data Path : Z:\HPCHEM1\BNA E\Data\Be051816\
 Data File : BE091791.D
 Acq On : 18 May 2016 13:49
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 18 18:03:23 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	-0.02
2	1,4-Dioxane	0.752	0.578	23.1	67	-0.02
3	n-Nitrosodimethylamine	0.908	0.746	17.8	74	0.00
4 S	2-Fluorophenol	1.228	1.094	10.9	82	-0.02
5 S	Phenol-d6	1.629	1.435	11.9	83	0.00
6	bis(2-Chloroethyl)ether	1.433	1.359	5.2	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	91	-0.02
8 S	Nitrobenzene-d5	0.323	0.282	12.7	85	-0.01
9	Nitrobenzene	0.338	0.284	16.0	83	-0.01
10	Naphthalene	1.010	0.964	4.6	86	0.00
11	Hexachlorobutadiene	0.149	0.143	4.0	89	-0.02
12	2-Methylnaphthalene	0.649	0.616	5.1	86	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.01
14 S	2,4,6-Tribromophenol	0.157	0.156	0.6	102	-0.01
15 S	2-Fluorobiphenyl	1.414	1.318	6.8	88	-0.01
16	Acenaphthylene	1.923	1.894	1.5	105	0.00
17	Acenaphthene	1.247	1.115	10.6	84	-0.01
18	Fluorene	1.550	1.508	2.7	91	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	-0.01
20	4-Bromophenyl-phenylether	0.180	0.172	4.4	93	0.00
21	Hexachlorobenzene	0.180	0.175	2.8	95	0.00
22	Pentachlorophenol	0.085	0.072	15.3	97	0.00
23	Phenanthrene	1.136	1.129	0.6	95	-0.01
24	Anthracene	1.025	0.978	4.6	95	-0.01
25	Fluoranthene	1.188	1.273	-7.2	104	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	121	-0.02
27	Benzydine	0.286	0.195	31.8#	102	0.00
28	Pvrene	1.130	0.894	20.9	99	0.00
29 S	Terphenyl-d14	0.780	0.585	25.0#	93	-0.02
30	Benzo(a)anthracene	1.221	1.015	16.9	103	-0.02
31	3,3'-Dichlorobenzidine	0.623	0.583	6.4	131	-0.02
32	Chrysene	1.269	0.953	24.9	86	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.417	-12.7	177#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.122	0.7	118	-0.03
35 I	Perylene-d12	1.000	1.000	0.0	123	-0.02
36	Benzo(b)fluoranthene	1.172	1.022	12.8	111	-0.02
37	Benzo(k)fluoranthene	1.191	1.054	11.5	114	-0.02
38 C	Benzo(a)pyrene	1.103	0.987	10.5	117	-0.02
39	Dibenzo(a,h)anthracene	1.045	0.958	8.3	116	-0.05
40	Benzo(a,h,i)perylene	1.062	0.988	7.0	119	-0.05

Data Path : Z:\HPCHEM1\BNA E\Data\Be051816\
Data File : BE091791.D
Acq On : 18 May 2016 13:49
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 18 18:03:23 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				