

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051816\
 Data File : BE091798.D
 Acq On : 18 May 2016 20:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 19 17:00:51 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	-0.02
2	1,4-Dioxane	0.752	0.602	19.9	61	0.00
3	n-Nitrosodimethylamine	0.908	0.781	14.0	68	0.00
4 S	2-Fluorophenol	1.228	1.258	-2.4	82	0.00
5 S	Phenol-d6	1.629	1.759	-8.0	89	0.00
6	bis(2-Chloroethyl)ether	1.433	1.384	3.4	74	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.323	0.321	0.6	92	0.00
9	Nitrobenzene	0.338	0.322	4.7	89	-0.01
10	Naphthalene	1.010	1.028	-1.8	87	0.00
11	Hexachlorobutadiene	0.149	0.153	-2.7	90	-0.02
12	2-Methylnaphthalene	0.649	0.689	-6.2	92	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	-0.01
14 S	2,4,6-Tribromophenol	0.157	0.177	-12.7	120	-0.01
15 S	2-Fluorobiphenyl	1.414	1.318	6.8	92	-0.01
16	Acenaphthylene	1.923	2.131	-10.8	122	0.00
17	Acenaphthene	1.247	1.224	1.8	96	-0.01
18	Fluorene	1.550	1.593	-2.8	99	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	-0.01
20	4-Bromophenyl-phenylether	0.180	0.169	6.1	102	0.00
21	Hexachlorobenzene	0.180	0.169	6.1	103	0.00
22	Pentachlorophenol	0.085	0.069	18.8	104	0.00
23	Phenanthrene	1.136	1.024	9.9	96	-0.01
24	Anthracene	1.025	0.962	6.1	104	-0.01
25	Fluoranthene	1.188	1.170	1.5	106	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	102	-0.02
27	Benzydine	0.286	0.191	33.2#	84	0.00
28	Pvrene	1.130	1.100	2.7	103	0.00
29 S	Terphenyl-d14	0.780	0.697	10.6	93	-0.02
30	Benzo(a)anthracene	1.221	1.124	7.9	96	-0.02
31	3,3'-Dichlorobenzidine	0.623	0.848	-36.1#	160#	-0.02
32	Chrysene	1.269	1.115	12.1	85	0.00
33	Bis(2-ethylhexyl)phthalate	0.370	0.775	-109.5#	278#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.130	1.240	-9.7	110	-0.03
35 I	Perylene-d12	1.000	1.000	0.0	101	-0.02
36	Benzo(b)fluoranthene	1.172	1.169	0.3	103	-0.02
37	Benzo(k)fluoranthene	1.191	1.234	-3.6	109	-0.01
38 C	Benzo(a)pyrene	1.103	1.153	-4.5	111	-0.02
39	Dibenzo(a,h)anthracene	1.045	1.113	-6.5	110	-0.05
40	Benzo(a,h,i)perylene	1.062	1.135	-6.9	111	-0.05

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051816\
Data File : BE091798.D
Acq On : 18 May 2016 20:59
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 19 17:00:51 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				