

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051816\
 Data File : BE091798.D
 Acq On : 18 May 2016 20:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 19 17:00:51 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 26 16:09:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	77	-0.02
2	1,4-Dioxane	0.400	0.319	20.3	61	0.00
3	n-Nitrosodimethylamine	0.400	0.344	14.0	68	0.00
4 S	2-Fluorophenol	0.400	0.410	-2.5	82	0.00
5 S	Phenol-d6	0.400	0.432	-8.0	89	0.00
6	bis(2-Chloroethyl)ether	0.400	0.386	3.5	74	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	92	0.00
9	Nitrobenzene	0.400	0.382	4.5	89	-0.01
10	Naphthalene	0.400	0.407	-1.7	87	0.00
11	Hexachlorobutadiene	0.400	0.409	-2.2	90	-0.02
12	2-Methylnaphthalene	0.400	0.425	-6.2	92	-0.01
13 I	Acenaphthene-d10	5.000	5.000	0.0	97	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.535	-33.8#	120	-0.01
15 S	2-Fluorobiphenyl	0.400	0.373	6.8	92	-0.01
16	Acenaphthylene	0.400	0.443	-10.7	122	0.00
17	Acenaphthene	0.400	0.393	1.8	96	-0.01
18	Fluorene	0.400	0.411	-2.7	99	-0.01
19 I	Phenanthrene-d10	5.000	5.000	0.0	109	-0.01
20	4-Bromophenyl-phenylether	0.400	0.376	6.0	102	0.00
21	Hexachlorobenzene	0.400	0.377	5.8	103	0.00
22	Pentachlorophenol	0.400	0.376	6.0	104	0.00
23	Phenanthrene	0.400	0.361	9.8	96	-0.01
24	Anthracene	0.400	0.375	6.3	104	-0.01
25	Fluoranthene	0.400	0.394	1.5	106	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	102	-0.02
27	Benzydine	0.400	0.350	12.5	84	0.00
28	Pvrene	0.400	0.390	2.5	103	0.00
29 S	Terphenyl-d14	0.400	0.358	10.5	93	-0.02
30	Benzo(a)anthracene	0.400	0.368	8.0	96	-0.02
31	3,3'-Dichlorobenzidine	0.400	0.545	-36.3#	160	-0.02
32	Chrysene	0.400	0.351	12.3	85	0.00
33	Bis(2-ethylhexyl)phthalate	0.400	0.838	-109.5#	278	-0.02
34	Indeno(1,2,3-cd)pyrene	0.400	0.439	-9.7	110	-0.03
35 I	Perylene-d12	5.000	5.000	0.0	101	-0.02
36	Benzo(b)fluoranthene	0.400	0.399	0.3	103	-0.02
37	Benzo(k)fluoranthene	0.400	0.414	-3.5	109	-0.01
38 C	Benzo(a)pyrene	0.400	0.418	-4.5	111	-0.02
39	Dibenzo(a,h)anthracene	0.400	0.426	-6.5	110	-0.05
40	Benzo(a,h,i)perylene	0.400	0.427	-6.7	111	-0.05

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051816\
Data File : BE091798.D
Acq On : 18 May 2016 20:59
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 19 17:00:51 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE042616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 26 16:09:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				