

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051916\
 Data File : BE091815.D
 Acq On : 19 May 2016 18:13
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 20 05:09:43 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 19 15:36:49 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	-0.02
2	1,4-Dioxane	0.649	0.634	2.3	94	0.00
3	n-Nitrosodimethylamine	0.798	0.687	13.9	96	0.00
4 S	2-Fluorophenol	1.307	1.258	3.7	104	0.00
5 S	Phenol-d6	1.840	1.715	6.8	106	0.00
6	bis(2-Chloroethyl)ether	1.477	1.337	9.5	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	-0.02
8 S	Nitrobenzene-d5	0.331	0.303	8.5	107	0.00
9	Nitrobenzene	0.341	0.307	10.0	109	-0.01
10	Naphthalene	1.106	1.072	3.1	106	0.00
11	Hexachlorobutadiene	0.164	0.158	3.7	105	-0.02
12	2-Methylnaphthalene	0.744	0.712	4.3	106	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	-0.01
14 S	2,4,6-Tribromophenol	0.194	0.165	14.9	102	-0.01
15 S	2-Fluorobiphenyl	1.414	1.356	4.1	105	-0.01
16	Acenaphthylene	2.230	2.131	4.4	105	0.00
17	Acenaphthene	1.341	1.236	7.8	102	-0.01
18	Fluorene	1.738	1.616	7.0	103	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	-0.01
20	4-Bromophenyl-phenylether	0.187	0.189	-1.1	104	0.00
21	Hexachlorobenzene	0.195	0.200	-2.6	104	0.00
22	Pentachlorophenol	0.080	0.082	-2.5	113	0.00
23	Phenanthrene	1.190	1.230	-3.4	103	-0.01
24	Anthracene	1.088	1.084	0.4	103	-0.01
25	Fluoranthene	1.329	1.328	0.1	100	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	101	-0.02
27	Benzydine	0.338	0.216	36.1#	84	0.00
28	Pyrene	1.097	1.026	6.5	99	0.00
29 S	Terphenyl-d14	0.708	0.690	2.5	106	-0.02
30	Benzo(a)anthracene	1.206	1.171	2.9	105	-0.02
31	3,3'-Dichlorobenzidine	0.767	0.706	8.0	106	-0.02
32	Chrysene	1.101	1.041	5.4	103	0.00
33	Bis(2-ethylhexyl)phthalate	0.540	0.535	0.9	126	-0.02
34	Indeno(1,2,3-cd)pyrene	1.236	1.146	7.3	100	-0.03
35 I	Perylene-d12	1.000	1.000	0.0	99	-0.02
36	Benzo(b)fluoranthene	1.251	1.205	3.7	101	-0.02
37	Benzo(k)fluoranthene	1.281	1.243	3.0	102	-0.02
38 C	Benzo(a)pyrene	1.194	1.131	5.3	100	-0.02
39	Dibenzo(a,h)anthracene	1.166	1.095	6.1	100	-0.05
40	Benzo(g,h,i)perylene	1.203	1.139	5.3	99	-0.05

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051916\
Data File : BE091815.D
Acq On : 19 May 2016 18:13
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 20 05:09:43 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 19 15:36:49 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				