

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051916\
 Data File : BE091829.D
 Acq On : 20 May 2016 2:58
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 20 15:12:59 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 19 15:36:49 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	-0.02
2	1,4-Dioxane	0.649	0.691	-6.5	101	0.00
3	n-Nitrosodimethylamine	0.798	0.782	2.0	108	0.00
4 S	2-Fluorophenol	1.307	1.499	-14.7	123	0.00
5 S	Phenol-d6	1.840	2.113	-14.8	129	0.02
6	bis(2-Chloroethyl)ether	1.477	1.443	2.3	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.331	0.345	-4.2	115	0.00
9	Nitrobenzene	0.341	0.349	-2.3	117	-0.01
10	Naphthalene	1.106	1.063	3.9	99	0.00
11	Hexachlorobutadiene	0.164	0.156	4.9	98	-0.02
12	2-Methylnaphthalene	0.744	0.710	4.6	100	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.01
14 S	2,4,6-Tribromophenol	0.194	0.211	-8.8	119	0.00
15 S	2-Fluorobiphenyl	1.414	1.369	3.2	96	-0.01
16	Acenaphthylene	2.230	2.259	-1.3	101	0.00
17	Acenaphthene	1.341	1.316	1.9	99	-0.01
18	Fluorene	1.738	1.696	2.4	98	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	-0.01
20	4-Bromophenyl-phenylether	0.187	0.173	7.5	97	0.00
21	Hexachlorobenzene	0.195	0.176	9.7	94	0.00
22	Pentachlorophenol	0.080	0.070	12.5	99	0.00
23	Phenanthrene	1.190	1.095	8.0	93	-0.01
24	Anthracene	1.088	1.030	5.3	100	-0.01
25	Fluoranthene	1.329	1.342	-1.0	103	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	93	-0.02
27	Benzydine	0.338	0.211	37.6#	76	0.00
28	Pyrene	1.097	1.146	-4.5	102	0.00
29 S	Terphenyl-d14	0.708	0.739	-4.4	104	-0.02
30	Benzo(a)anthracene	1.206	1.267	-5.1	105	-0.02
31	3,3'-Dichlorobenzidine	0.767	0.978	-27.5#	135	-0.02
32	Chrysene	1.101	1.037	5.8	94	0.00
33	Bis(2-ethylhexyl)phthalate	0.540	0.633	-17.2	137	-0.02
34	Indeno(1,2,3-cd)pyrene	1.236	1.270	-2.8	101	-0.03
35 I	Perylene-d12	1.000	1.000	0.0	96	-0.02
36	Benzo(b)fluoranthene	1.251	1.224	2.2	100	-0.02
37	Benzo(k)fluoranthene	1.281	1.283	-0.2	102	-0.02
38 C	Benzo(a)pyrene	1.194	1.206	-1.0	103	-0.02
39	Dibenzo(a,h)anthracene	1.166	1.170	-0.3	103	-0.05
40	Benzo(g,h,i)perylene	1.203	1.197	0.5	101	-0.05

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051916\
Data File : BE091829.D
Acq On : 20 May 2016 2:58
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 20 15:12:59 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 19 15:36:49 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				