

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051916\
 Data File : BE091839.D
 Acq On : 20 May 2016 16:06
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 20 17:23:27 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 19 15:36:49 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	140	0.00
2	1,4-Dioxane	0.649	0.545	16.0	111	0.00
3	n-Nitrosodimethylamine	0.798	0.726	9.0	139	0.00
4 S	2-Fluorophenol	1.307	1.369	-4.7	156#	0.00
5 S	Phenol-d6	1.840	1.957	-6.4	166#	0.02
6	bis(2-Chloroethyl)ether	1.477	1.436	2.8	149	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	141	0.00
8 S	Nitrobenzene-d5	0.331	0.339	-2.4	162#	0.00
9	Nitrobenzene	0.341	0.336	1.5	161#	0.00
10	Naphthalene	1.106	1.061	4.1	141	0.00
11	Hexachlorobutadiene	0.164	0.162	1.2	146	-0.02
12	2-Methylnaphthalene	0.744	0.687	7.7	138	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	122	-0.01
14 S	2,4,6-Tribromophenol	0.194	0.184	5.2	133	0.00
15 S	2-Fluorobiphenyl	1.414	1.412	0.1	129	-0.01
16	Acenaphthylene	2.230	2.257	-1.2	130	0.00
17	Acenaphthene	1.341	1.309	2.4	127	-0.01
18	Fluorene	1.738	1.599	8.0	120	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	119	-0.01
20	4-Bromophenyl-phenylether	0.187	0.183	2.1	119	0.00
21	Hexachlorobenzene	0.195	0.181	7.2	112	0.00
22	Pentachlorophenol	0.080	0.051	36.3#	83	0.00
23	Phenanthrene	1.190	1.088	8.6	108	0.00
24	Anthracene	1.088	1.018	6.4	115	-0.01
25	Fluoranthene	1.329	1.254	5.6	112	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	94	-0.02
27	Benzydine	0.338	0.479	-41.7#	175#	0.00
28	Pyrene	1.097	1.216	-10.8	110	0.00
29 S	Terphenyl-d14	0.708	0.771	-8.9	110	-0.02
30	Benzo(a)anthracene	1.206	1.215	-0.7	102	-0.02
31	3,3'-Dichlorobenzidine	0.767	1.069	-39.4#	150#	-0.02
32	Chrysene	1.101	1.141	-3.6	105	0.00
33	Bis(2-ethylhexyl)phthalate	0.540	0.767	-42.0#	168#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.236	1.360	-10.0	110	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	101	-0.02
36	Benzo(b)fluoranthene	1.251	1.242	0.7	106	-0.02
37	Benzo(k)fluoranthene	1.281	1.305	-1.9	108	-0.01
38 C	Benzo(a)pyrene	1.194	1.231	-3.1	110	-0.02
39	Dibenzo(a,h)anthracene	1.166	1.195	-2.5	110	-0.05
40	Benzo(g,h,i)perylene	1.203	1.213	-0.8	107	-0.05

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE051916\
Data File : BE091839.D
Acq On : 20 May 2016 16:06
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 20 17:23:27 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 19 15:36:49 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				